

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071525\
 Data File : P0112308.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2025 23:44
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 16 01:52:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 09 10:50:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.671	3.664	676.7E6	407.8E6	68.386	64.012
2) SA Decachlor...	8.698	8.645	371.3E6	90696692	54.308	52.625
Target Compounds						
26) L6 AR-1254-1	5.563	5.538	381.8E6	218.3E6	620.700	572.619
27) L6 AR-1254-2	5.713	5.686	327.1E6	185.7E6	610.567	559.534
28) L6 AR-1254-3	6.117	6.087	466.1E6	261.4E6	555.126	542.314
29) L6 AR-1254-4	6.346	6.314	322.8E6	154.8E6	671.067	593.976
30) L6 AR-1254-5	6.765	6.731	406.7E6	190.3E6	525.454	533.931



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO071525\
Data File : PO112308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jul 2025 23:44
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 01:52:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO070825.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 09 10:50:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.671 min
Delta R.T.: -0.002 min
Response: 676661382
Conc: 68.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.664 min
Delta R.T.: -0.020 min
Response: 407823294
Conc: 64.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.698 min
Delta R.T.: -0.005 min
Response: 371303841
Conc: 54.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.645 min
Delta R.T.: -0.023 min
Response: 90696692
Conc: 52.62 ng/ml



#26 AR-1254-1


R.T.: 5.563 min
Delta R.T.: -0.002 min
Response: 381800497
Conc: 620.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.538 min
Delta R.T.: -0.004 min
Response: 218264051
Conc: 572.62 ng/ml




#27 AR-1254-2

R.T.: 5.713 min
Delta R.T.: -0.002 min
Response: 327072338
Conc: 610.57 ng/ml



#27 AR-1254-2


R.T.: 5.686 min
Delta R.T.: -0.003 min
Response: 185659701
Conc: 559.53 ng/ml



#28 AR-1254-3


R.T.: 6.117 min
Delta R.T.: -0.002 min
Response: 466071606
Conc: 555.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.087 min
Delta R.T.: -0.003 min
Response: 261418656
Conc: 542.31 ng/ml




#29 AR-1254-4

R.T.: 6.346 min
Delta R.T.: -0.002 min
Response: 322795687
Conc: 671.07 ng/ml



#29 AR-1254-4


R.T.: 6.314 min
Delta R.T.: -0.002 min
Response: 154846466
Conc: 593.98 ng/ml



#30 AR-1254-5

R.T.: 6.765 min
Delta R.T.: -0.003 min
Response: 406674282
Conc: 525.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.731 min
Delta R.T.: -0.003 min
Response: 190269567
Conc: 533.93 ng/ml