

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0071618\
 Data File : P0046788.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Jul 2018 2:38
 Operator : SM/SJ
 Sample : J3819-03
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AU-05-071318-B

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 17 07:43:42 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jul 09 15:27:34 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.195	3.519	4845368	5721735	20.275	26.938 #
2)	SA Decachlor...	9.718	8.445	2917700	2485117	15.308	12.811

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0071618\
Data File : P0046788.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Jul 2018 2:38
Operator : SM/SJ
Sample : J3819-03
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AU-05-071318-B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 07:43:42 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070718.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 09 15:27:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.195 min
Delta R.T.: -0.018 min
Response: 4845368
Conc: 20.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AU-05-071318-B




#1 Tetrachloro-m-xylene

R.T.: 3.519 min
Delta R.T.: -0.020 min
Response: 5721735
Conc: 26.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.718 min
Delta R.T.: -0.030 min
Response: 2917700
Conc: 15.31 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.445 min
Delta R.T.: -0.028 min
Response: 2485117
Conc: 12.81 ng/ml