

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071620\
 Data File : P0069757.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Jul 2020 23:12
 Operator : DD\AJ
 Sample : L3332-02
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S3-05-B-01

Manual Integrations
 APPROVED

Ankita
 7/17/2020 11:39:55 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 17 01:43:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 08 16:12:41 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.389	3.565	1146780	1297951	30.715m	32.615
2) SA Decachlor...	10.044	8.770	1938775	1719244	32.224	30.293
Target Compounds						
26) L6 AR-1254-1	6.590	5.652	612928	577304	307.914m	221.776 #
27) L6 AR-1254-2	6.821	5.814	1272119	633335	396.269	274.340 #
28) L6 AR-1254-3	7.197	6.225	1624108	1134435	462.434	302.072 #
29) L6 AR-1254-4	7.492	6.466	1132723	784581	378.473	310.463
30) L6 AR-1254-5	7.914	6.891	1810425	6437974	601.711	1793.832 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071620\
Data File : P0069757.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Jul 2020 23:12
Operator : DD\AJ
Sample : L3332-02
Misc :
ALS Vial : 47 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
S3-05-B-01


Manual Integrations
APPROVED

Ankita
7/17/2020 11:39:55 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 01:43:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070820.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 08 16:12:41 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.389 min
Delta R.T.: -0.003 min
Response: 1146780
Conc: 30.72 ng/ml m

Instrument :
ECD_O
ClientSampleId :
S3-05-B-01


Manual Integrations
APPROVED

Ankita
7/17/2020 11:39:55 AM




#1 Tetrachloro-m-xylene

R.T.: 3.565 min
Delta R.T.: 0.003 min
Response: 1297951
Conc: 32.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.044 min
Delta R.T.: -0.009 min
Response: 1938775
Conc: 32.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.770 min
Delta R.T.: -0.004 min
Response: 1719244
Conc: 30.29 ng/ml




#26 AR-1254-1

R.T.: 6.590 min
Delta R.T.: -0.006 min
Response: 612928
Conc: 307.91 ng/ml m

Instrument :
ECD_O
ClientSampleId :
S3-05-B-01


Manual Integrations
APPROVED

Ankita
7/17/2020 11:39:55 AM




#26 AR-1254-1

R.T.: 5.652 min
Delta R.T.: -0.002 min
Response: 577304
Conc: 221.78 ng/ml




#27 AR-1254-2

R.T.: 6.821 min
Delta R.T.: -0.004 min
Response: 1272119
Conc: 396.27 ng/ml



#27 AR-1254-2

R.T.: 5.814 min
Delta R.T.: -0.001 min
Response: 633335
Conc: 274.34 ng/ml




#28 AR-1254-3

R.T.: 7.197 min
Delta R.T.: -0.005 min
Response: 1624108
Conc: 462.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
S3-05-B-01


Manual Integrations
APPROVED

Ankita
7/17/2020 11:39:55 AM




#28 AR-1254-3

R.T.: 6.225 min
Delta R.T.: -0.002 min
Response: 1134435
Conc: 302.07 ng/ml




#29 AR-1254-4

R.T.: 7.492 min
Delta R.T.: -0.007 min
Response: 1132723
Conc: 378.47 ng/ml



#29 AR-1254-4

R.T.: 6.466 min
Delta R.T.: -0.001 min
Response: 784581
Conc: 310.46 ng/ml




#30 AR-1254-5

R.T.: 7.914 min
Delta R.T.: -0.006 min
Response: 1810425
Conc: 601.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
S3-05-B-01

Manual Integrations
APPROVED

Ankita
7/17/2020 11:39:55 AM



#30 AR-1254-5

R.T.: 6.891 min
Delta R.T.: 0.000 min
Response: 6437974
Conc: 1793.83 ng/ml