

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071720\
 Data File : P0069778.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Jul 2020 14:28
 Operator : DD\AJ
 Sample : PB130271BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB130271BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 17 17:56:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 08 16:12:41 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.388	3.563	1183593	1224431	31.701	30.767
2)	SA Decachlor...	10.043	8.769	1187926	1059979	19.744	18.677

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071720\
Data File : P0069778.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Jul 2020 14:28
Operator : DD\AJ
Sample : PB130271BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB130271BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 17:56:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070820.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 08 16:12:41 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.388 min
Delta R.T.: -0.004 min
Response: 1183593
Conc: 31.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB130271BL




#1 Tetrachloro-m-xylene

R.T.: 3.563 min
Delta R.T.: 0.002 min
Response: 1224431
Conc: 30.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.043 min
Delta R.T.: -0.010 min
Response: 1187926
Conc: 19.74 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.769 min
Delta R.T.: -0.005 min
Response: 1059979
Conc: 18.68 ng/ml