

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071822\
 Data File : P0088131.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jul 2022 10:44
 Operator : YP/AJ
 Sample : N3736-01DL 5X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PIPE-COATING-LOCATION-B-ASTORIA-YARDDL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 18 12:09:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 15 03:57:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.445	3.603	12039749	3503260	4.719	3.713
2) SA Decachlor...	10.296	8.606	9883458	4698846	4.654	4.496
Target Compounds						
31) L7 AR-1260-1	7.238	6.165	43177934	28136872	371.082m	519.149m#
32) L7 AR-1260-2	7.503	6.353	70813108	50897109	524.101m	783.849 #
33) L7 AR-1260-3	7.856	6.507	27245182	23235109	297.597m	373.542 #
34) L7 AR-1260-4	8.082	6.979	32235476	19146106	298.413	404.778m#
35) L7 AR-1260-5	8.412	7.221	51602474	25707960	251.472	224.686m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071822\
Data File : P0088131.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jul 2022 10:44
Operator : YP/AJ
Sample : N3736-01DL 5X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PIPE-COATING-LOCATION-B-ASTORIA-YARDDL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 18 12:09:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 15 03:57:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

