

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071824\
 Data File : P0104834.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jul 2024 15:39
 Operator : YP/AJ
 Sample : P3294-05
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HR-4-071724

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/19/2024
 Supervised By :Ankita Jodhani 07/19/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 19 02:16:52 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 02 16:29:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.427	3.542	90122830	34709822	10.066m	10.639
2) SA Decachlor...	10.216	8.438	89561624	24072581	14.604m	11.479
Target Compounds						
31) L7 AR-1260-1	7.211	6.054	10781369	4319981	29.865m	26.978m
32) L7 AR-1260-2	7.463	6.240	46303636	5860805	109.748m	30.508m#
33) L7 AR-1260-3	7.835	6.390	11254439	3726608	38.094m	20.952 #
34) L7 AR-1260-4	8.062	6.857	11179409	3016838	31.554m	22.519 #
35) L7 AR-1260-5	8.383	7.096	22073473	7153294	33.868m	23.373 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO071824\
Data File : PO104834.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jul 2024 15:39
Operator : YP/AJ
Sample : P3294-05
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
HR-4-071724


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/19/2024
Supervised By :Ankita Jodhani 07/19/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 19 02:16:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO070224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 02 16:29:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.427 min
Delta R.T.: 0.002 min
Response: 90122830
Conc: 10.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
HR-4-071724


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/19/2024
Supervised By :Ankita Jodhani 07/19/2024




#1 Tetrachloro-m-xylene

R.T.: 3.542 min
Delta R.T.: 0.000 min
Response: 34709822
Conc: 10.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.216 min
Delta R.T.: 0.011 min
Response: 89561624
Conc: 14.60 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.438 min
Delta R.T.: -0.002 min
Response: 24072581
Conc: 11.48 ng/ml




#31 AR-1260-1

R.T.: 7.211 min
Delta R.T.: 0.006 min
Response: 10781369
Conc: 29.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
HR-4-071724


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/19/2024
Supervised By :Ankita Jodhani 07/19/2024




#31 AR-1260-1

R.T.: 6.054 min
Delta R.T.: -0.002 min
Response: 4319981
Conc: 26.98 ng/ml m




#32 AR-1260-2

R.T.: 7.463 min
Delta R.T.: 0.000 min
Response: 46303636
Conc: 109.75 ng/ml m



#32 AR-1260-2

R.T.: 6.240 min
Delta R.T.: -0.001 min
Response: 5860805
Conc: 30.51 ng/ml m




#33 AR-1260-3

R.T.: 7.835 min
Delta R.T.: 0.009 min
Response: 11254439
Conc: 38.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
HR-4-071724


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/19/2024
Supervised By :Ankita Jodhani 07/19/2024




#33 AR-1260-3

R.T.: 6.390 min
Delta R.T.: -0.002 min
Response: 3726608
Conc: 20.95 ng/ml




#34 AR-1260-4

R.T.: 8.062 min
Delta R.T.: 0.008 min
Response: 11179409
Conc: 31.55 ng/ml m



#34 AR-1260-4

R.T.: 6.857 min
Delta R.T.: -0.002 min
Response: 3016838
Conc: 22.52 ng/ml




#35 AR-1260-5

R.T.: 8.383 min
Delta R.T.: 0.008 min
Response: 22073473
Conc: 33.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
HR-4-071724

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/19/2024
Supervised By :Ankita Jodhani 07/19/2024



#35 AR-1260-5

R.T.: 7.096 min
Delta R.T.: -0.001 min
Response: 7153294
Conc: 23.37 ng/ml