

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0071918\
 Data File : P0046911.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Jul 2018 4:58
 Operator : SM/SJ
 Sample : J4031-03
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 PLANTER-DOC-3

Manual Integrations
 APPROVED

Sohil
 7/19/2018 5:38:19 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 19 16:32:34 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0071918.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 19 01:07:21 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.191	3.516	6076251	6371982	21.144	20.875
2) SA Decachlor...	9.713	8.438	3911518	4096598	17.784	16.731
Target Compounds						
31) L7 AR-1260-1	6.924	6.046	3871184	5419421	326.537	339.449m
32) L7 AR-1260-2	7.175	6.233	7147875	5731440	511.248	272.490m#
33) L7 AR-1260-3	7.459	6.558	7295319	13202838	452.976	597.701m#
34) L7 AR-1260-4	8.062	7.092	3026077	4482709	132.941	121.244m
35) L7 AR-1260-5	8.360	7.437	2214290	3333984	156.300	131.426

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071918\
Data File : P0046911.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Jul 2018 4:58
Operator : SM/SJ
Sample : J4031-03
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PLANTER-DOC-3

Manual Integrations
APPROVED

Sohil
7/19/2018 5:38:19 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 19 16:32:34 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071918.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 19 01:07:21 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

