

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0071918.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jul 19 01:07:21 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0046876.D 1000 =P0046874.D 750 =P0046875.D
 250 =P0046877.D 50 =P0046878.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		3.023	2.556	2.800	3.051	2.939	2.874	E5	7.06
2) SA	Decachlorobiphenyl		2.216	1.933	2.078	2.387	2.383	2.199	E5	8.96
3) L1	AR-1016-1		6.864	5.569	6.084	6.851	6.776	6.429	E3	9.03
4) L1	AR-1016-2		8.398	7.191	7.836	8.704	8.333	8.092	E3	7.32
5) L1	AR-1016-3		7.046	6.020	6.608	7.313	7.168	6.831	E3	7.68
6) L1	AR-1016-4		6.734	5.778	6.572	6.624	7.194	6.581	E3	7.77
7) L1	AR-1016-5		7.511	6.373	7.074	7.448	7.596	7.201	E3	7.00
8) L2	AR-1221-1		3.698					3.698	E3	0.00
9) L2	AR-1221-2		2.622					2.622	E3	0.00
10) L2	AR-1221-3		8.443					8.443	E3	0.00
11) L3	AR-1232-1		1.836					1.836	E3	0.00
12) L3	AR-1232-2		3.866					3.866	E3	0.00
13) L3	AR-1232-3		5.544					5.544	E3	0.00
14) L3	AR-1232-4		3.440					3.440	E3	0.00
15) L3	AR-1232-5		2.794					2.794	E3	0.00
16) L4	AR-1242-1		1.139					1.139	E4	0.00
17) L4	AR-1242-2		7.038					7.038	E3	0.00
18) L4	AR-1242-3		5.895					5.895	E3	0.00
19) L4	AR-1242-4		5.709					5.709	E3	0.00
20) L4	AR-1242-5		6.541					6.541	E3	0.00
21) L5	AR-1248-1		9.038					9.038	E3	0.00
22) L5	AR-1248-2		8.145					8.145	E3	0.00
23) L5	AR-1248-3		1.055					1.055	E4	0.00
24) L5	AR-1248-4		1.018					1.018	E4	0.00
25) L5	AR-1248-5		6.956					6.956	E3	0.00
26) L6	AR-1254-1		1.511					1.511	E4	0.00
27) L6	AR-1254-2		9.524					9.524	E3	0.00
28) L6	AR-1254-3		1.636					1.636	E4	0.00
29) L6	AR-1254-4		1.226					1.226	E4	0.00
30) L6	AR-1254-5		9.264					9.264	E3	0.00
31) L7	AR-1260-1		1.190	1.051	1.104	1.259	1.324	1.186	E4	9.37
32) L7	AR-1260-2		1.428	1.232	1.325	1.514	1.492	1.398	E4	8.48
33) L7	AR-1260-3		1.661	1.450	1.553	1.763	1.626	1.611	E4	7.28
34) L7	AR-1260-4		2.277	2.005	2.137	2.388	2.574	2.276	E4	9.68
35) L7	AR-1260-5		1.436	1.259	1.347	1.518	1.523	1.417	E4	8.02
36) L8	AR-1262-1		1.029					1.029	E4	0.00
37) L8	AR-1262-2		1.195					1.195	E4	0.00
38) L8	AR-1262-3		1.548					1.548	E4	0.00
39) L8	AR-1262-4		1.451					1.451	E4	0.00
40) L8	AR-1262-5		2.715					2.715	E4	0.00
41) L9	AR-1268-1		2.792					2.792	E4	0.00
42) L9	AR-1268-2		2.603					2.603	E4	0.00
43) L9	AR-1268-3		2.169					2.169	E4	0.00
44) L9	AR-1268-4		9.823					9.823	E3	0.00
45) L9	AR-1268-5		6.650					6.650	E4	0.00

Signal #2 Calibration Files

500 =P0046876.D 1000 =P0046874.D 750 =P0046875.D
 250 =P0046877.D 50 =P0046878.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		3.183	2.788	3.027	3.216	3.048	3.052	E5	5.54
2) SA	Decachlorobiphenyl		2.510	2.231	2.393	2.631	2.478	2.449	E5	6.08

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0071918.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jul 19 01:07:21 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0046876.D 1000 =P0046874.D 750 =P0046875.D
 250 =P0046877.D 50 =P0046878.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	7.602	6.639	7.114	7.738	7.198	7.258	E3	5.99
4)	L1	AR-1016-2	8.225	7.234	7.748	8.478	8.284	7.994	E3	6.29
5)	L1	AR-1016-3	6.014	5.281	5.704	6.240	5.650	5.778	E3	6.35
6)	L1	AR-1016-4	7.431	6.527	7.011	7.637	6.327	6.987	E3	8.06
7)	L1	AR-1016-5	8.110	7.263	7.643	8.288	7.638	7.788	E3	5.27
8)	L2	AR-1221-1	3.993					3.993	E3	0.00
9)	L2	AR-1221-2	2.876					2.876	E3	0.00
10)	L2	AR-1221-3	8.999					8.999	E3	0.00
11)	L3	AR-1232-1	1.888					1.888	E3	0.00
12)	L3	AR-1232-2	4.447					4.447	E3	0.00
13)	L3	AR-1232-3	6.198					6.198	E3	0.00
14)	L3	AR-1232-4	4.250					4.250	E3	0.00
15)	L3	AR-1232-5	2.196					2.196	E3	0.00
16)	L4	AR-1242-1	1.232					1.232	E4	0.00
17)	L4	AR-1242-2	6.850					6.850	E3	0.00
18)	L4	AR-1242-3	6.292					6.292	E3	0.00
19)	L4	AR-1242-4	7.072					7.072	E3	0.00
20)	L4	AR-1242-5	6.424					6.424	E3	0.00
21)	L5	AR-1248-1	1.091					1.091	E4	0.00
22)	L5	AR-1248-2	8.222					8.222	E3	0.00
23)	L5	AR-1248-3	1.618					1.618	E4	0.00
24)	L5	AR-1248-4	1.133					1.133	E4	0.00
25)	L5	AR-1248-5	7.621					7.621	E3	0.00
26)	L6	AR-1254-1	1.662					1.662	E4	0.00
27)	L6	AR-1254-2	1.423					1.423	E4	0.00
28)	L6	AR-1254-3	2.395					2.395	E4	0.00
29)	L6	AR-1254-4	1.638					1.638	E4	0.00
30)	L6	AR-1254-5	1.136					1.136	E4	0.00
31)	L7	AR-1260-1	1.618	1.432	1.508	1.692	1.732	1.597	E4	7.86
32)	L7	AR-1260-2	2.116	1.866	1.972	2.218	2.346	2.103	E4	9.07
33)	L7	AR-1260-3	2.241	1.949	2.075	2.316	2.464	2.209	E4	9.15
34)	L7	AR-1260-4	3.791	3.374	3.544	3.822	3.955	3.697	E4	6.32
35)	L7	AR-1260-5	2.563	2.252	2.568	2.637	2.663	2.537	E4	6.51
36)	L8	AR-1262-1	1.766					1.766	E4	0.00
37)	L8	AR-1262-2	1.666					1.666	E4	0.00
38)	L8	AR-1262-3	2.518					2.518	E4	0.00
39)	L8	AR-1262-4	2.038					2.038	E4	0.00
40)	L8	AR-1262-5	4.507					4.507	E4	0.00
41)	L9	AR-1268-1	4.350					4.350	E4	0.00
42)	L9	AR-1268-2	4.017					4.017	E4	0.00
43)	L9	AR-1268-3	3.317					3.317	E4	0.00
44)	L9	AR-1268-4	1.360					1.360	E4	0.00
45)	L9	AR-1268-5	8.728					8.728	E4	0.00

(#) = Out of Range