

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071922\
 Data File : P0088180.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2022 05:34
 Operator : YP/AJ
 Sample : N3763-02
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TRE-1279-WATER

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 06:21:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 15 03:57:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.446	3.606	40654171	16446124	15.935	17.431
2) SA Decachlor...	10.292	8.604	22076534	11923653	10.395	11.410

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071922\
Data File : P0088180.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2022 05:34
Operator : YP/AJ
Sample : N3763-02
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TRE-1279-WATER

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 06:21:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 15 03:57:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.446 min
Delta R.T.: 0.000 min
Response: 40654171
Conc: 15.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
TRE-1279-WATER




#1 Tetrachloro-m-xylene

R.T.: 3.606 min
Delta R.T.: 0.000 min
Response: 16446124
Conc: 17.43 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.292 min
Delta R.T.: -0.007 min
Response: 22076534
Conc: 10.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.604 min
Delta R.T.: -0.005 min
Response: 11923653
Conc: 11.41 ng/ml