

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072018\
Data File : P0046951.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Jul 2018 19:06
Operator : SM/SJ
Sample : J4033-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
071618-TIPC-E-U-M

Manual Integrations
APPROVED

Sohil
7/20/2018 11:24:49 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 02:27:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0071918.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 19 01:07:21 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.197	3.522	3033110	3194384	10.555m	10.465m
2)	SA Decachlor...	9.723	8.446	3151770	3278887	14.330	13.391

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072018\
Data File : P0046951.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Jul 2018 19:06
Operator : SM/SJ
Sample : J4033-02
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
071618-TIPC-E-U-M


Manual Integrations
APPROVED

Sohil
7/20/2018 11:24:49 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 02:27:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071918.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 19 01:07:21 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.197 min
Delta R.T.: 0.004 min
Response: 3033110
Conc: 10.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
071618-TIPC-E-U-M


Manual Integrations
APPROVED

Sohil
7/20/2018 11:24:49 AM




#1 Tetrachloro-m-xylene

R.T.: 3.522 min
Delta R.T.: 0.005 min
Response: 3194384
Conc: 10.46 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.723 min
Delta R.T.: 0.006 min
Response: 3151770
Conc: 14.33 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.446 min
Delta R.T.: 0.005 min
Response: 3278887
Conc: 13.39 ng/ml