

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072019\
 Data File : P0058192.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jul 2019 2:00
 Operator : SM/AJ
 Sample : K3835-43
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-6(0-2)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 03:29:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 17 11:13:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.192	3.535	709927	724751	20.950	22.597
2) SA Decachlor...	9.701	8.437	712965	482535	13.936	14.078
Target Compounds						
12) L3 AR-1232-2	5.024	0.000	16052	0	34.709	N.D. #
28) L6 AR-1254-3	6.791	0.000	19107	0	5.416	N.D. #
30) L6 AR-1254-5	7.467	0.000	158987	0	54.165	N.D. #
32) L7 AR-1260-2	7.185	0.000	201460	0	57.300	N.D. #
33) L7 AR-1260-3	7.536	0.000	26087	0	8.885	N.D. #
34) L7 AR-1260-4	7.722	6.852	94532	39571	34.605	20.437 #
35) L7 AR-1260-5	0.000	7.100	0	16954	N.D.	3.684 #
36) L8 AR-1262-1	7.536	0.000	26087	0	4.722	N.D. #
37) L8 AR-1262-2	0.000	7.100	0	16954	N.D.	2.625 #
38) L8 AR-1262-3	0.000	7.375	0	11543	N.D.	4.188 #
39) L8 AR-1262-4	0.000	7.442	0	7795	N.D.	1.633 #
41) L9 AR-1268-1	0.000	7.375	0	11543	N.D.	1.479 #
42) L9 AR-1268-2	0.000	7.442	0	7795	N.D.	1.094 #
43) L9 AR-1268-3	0.000	7.639	0	5659	N.D.	0.940 #
45) L9 AR-1268-5	0.000	8.204	0	80467	N.D.	4.825 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072019\
Data File : P0058192.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jul 2019 2:00
Operator : SM/AJ
Sample : K3835-43
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
SB-6(0-2)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 03:29:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071719.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 17 11:13:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.192 min
Delta R.T.: -0.008 min
Response: 709927
Conc: 20.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-6(0-2)




#1 Tetrachloro-m-xylene

R.T.: 3.535 min
Delta R.T.: -0.004 min
Response: 724751
Conc: 22.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.701 min
Delta R.T.: -0.017 min
Response: 712965
Conc: 13.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.437 min
Delta R.T.: -0.008 min
Response: 482535
Conc: 14.08 ng/ml



#12 AR-1232-2


R.T.: 5.024 min
Delta R.T.: -0.067 min
Response: 16052
Conc: 34.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-6(0-2)




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 4.619 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 6.791 min
Delta R.T.: 0.014 min
Response: 19107
Conc: 5.42 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 5.932 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 7.467 min
Delta R.T.: -0.008 min
Response: 158987
Conc: 54.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-6(0-2)




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 6.571 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 7.185 min
Delta R.T.: -0.009 min
Response: 201460
Conc: 57.30 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T. : 6.247 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.536 min
Delta R.T.: -0.012 min
Response: 26087
Conc: 8.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-6(0-2)




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.397 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 7.722 min
Delta R.T.: -0.051 min
Response: 94532
Conc: 34.61 ng/ml



#34 AR-1260-4


R.T.: 6.852 min
Delta R.T.: -0.010 min
Response: 39571
Conc: 20.44 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 8.079 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SB-6(0-2)




#35 AR-1260-5

R.T.: 7.100 min
Delta R.T.: -0.005 min
Response: 16954
Conc: 3.68 ng/ml




#36 AR-1262-1

R.T.: 7.536 min
Delta R.T.: -0.012 min
Response: 26087
Conc: 4.72 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 6.606 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 8.080 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SB-6(0-2)




#37 AR-1262-2

R.T.: 7.100 min
Delta R.T.: -0.005 min
Response: 16954
Conc: 2.62 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.363 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.375 min
Delta R.T.: -0.006 min
Response: 11543
Conc: 4.19 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 8.447 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SB-6(0-2)




#39 AR-1262-4

R.T.: 7.442 min
Delta R.T.: -0.005 min
Response: 7795
Conc: 1.63 ng/ml




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T.: 8.361 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 7.375 min
Delta R.T.: -0.006 min
Response: 11543
Conc: 1.48 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T.: 8.449 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SB-6(0-2)




#42 AR-1268-2

R.T.: 7.442 min
Delta R.T.: -0.006 min
Response: 7795
Conc: 1.09 ng/ml




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T.: 8.653 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 7.639 min
Delta R.T.: -0.006 min
Response: 5659
Conc: 0.94 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.420 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SB-6(0-2)



#45 AR-1268-5

R.T.: 8.204 min
Delta R.T.: -0.006 min
Response: 80467
Conc: 4.83 ng/ml