

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072021\
 Data File : PP037447.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jul 2021 20:40
 Operator : DD\AJ
 Sample : PB137739BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 02:14:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 07 08:32:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.958	3.948	977425	567622	21.506	20.743
2)	SA Decachlor...	10.989	9.240	918205	507674	19.946	18.239
Target Compounds							
9)	L2 AR-1221-2	5.311	0.000	12934	0	31.926	N.D. #
24)	L5 AR-1248-4	7.196f	0.000	7303	0	3.864	N.D. #
26)	L6 AR-1254-1	7.196	0.000	7303	0	4.065	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP072021\
Data File : PP037447.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jul 2021 20:40
Operator : DD\AJ
Sample : PB137739BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 02:14:35 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070621.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 07 08:32:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.958 min
Delta R.T.: -0.002 min
Response: 977425
Conc: 21.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.948 min
Delta R.T.: -0.002 min
Response: 567622
Conc: 20.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.989 min
Delta R.T.: -0.003 min
Response: 918205
Conc: 19.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.240 min
Delta R.T.: -0.007 min
Response: 507674
Conc: 18.24 ng/ml



#9 AR-1221-2


R.T.: 5.311 min
Delta R.T.: 0.002 min
Response: 12934
Conc: 31.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 4.305 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 7.196 min
Delta R.T.: -0.024 min
Response: 7303
Conc: 3.86 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.692 min
Response: 0
Conc: N.D.



#26 AR-1254-1

R.T.: 7.196 min
Delta R.T.: 0.006 min
Response: 7303
Conc: 4.07 ng/ml

Instrument :
ECD_O
ClientSampleId :



#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 6.073 min
Response: 0
Conc: N.D.