

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072021\
 Data File : P0079683.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jul 2021 10:00
 Operator : DD\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 14:42:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 20 14:31:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.015	4.228	3693366	1230919	49.894	48.992
2)	SA Decachlor...	11.142	9.698	2680602	1131470	49.854	49.778
Target Compounds							
16)	L4 AR-1242-1	6.347	5.512	962277	368430	498.490	484.356
17)	L4 AR-1242-2	6.371	5.533	1326456	487025	488.763	480.682
18)	L4 AR-1242-3	6.437	5.731	832650	265565	493.655	480.995
19)	L4 AR-1242-4	6.545	5.827	677694	272405	496.952	493.199
20)	L4 AR-1242-5	7.332	6.406	684572	334442	497.547	487.097



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072021\
Data File : P0079683.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jul 2021 10:00
Operator : DD\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 14:42:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071921.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 20 14:31:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.015 min
Delta R.T.: 0.000 min
Response: 3693366
Conc: 49.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.228 min
Delta R.T.: -0.010 min
Response: 1230919
Conc: 48.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.142 min
Delta R.T.: 0.000 min
Response: 2680602
Conc: 49.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.698 min
Delta R.T.: -0.007 min
Response: 1131470
Conc: 49.78 ng/ml



#16 AR-1242-1


R.T.: 6.347 min
Delta R.T.: -0.002 min
Response: 962277
Conc: 498.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.512 min
Delta R.T.: -0.009 min
Response: 368430
Conc: 484.36 ng/ml




#17 AR-1242-2

R.T.: 6.371 min
Delta R.T.: -0.002 min
Response: 1326456
Conc: 488.76 ng/ml



#17 AR-1242-2


R.T.: 5.533 min
Delta R.T.: -0.010 min
Response: 487025
Conc: 480.68 ng/ml



#18 AR-1242-3


R.T.: 6.437 min
Delta R.T.: -0.002 min
Response: 832650
Conc: 493.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.731 min
Delta R.T.: -0.008 min
Response: 265565
Conc: 481.00 ng/ml




#19 AR-1242-4

R.T.: 6.545 min
Delta R.T.: -0.003 min
Response: 677694
Conc: 496.95 ng/ml



#19 AR-1242-4


R.T.: 5.827 min
Delta R.T.: -0.008 min
Response: 272405
Conc: 493.20 ng/ml



#20 AR-1242-5

R.T.: 7.332 min
Delta R.T.: -0.002 min
Response: 684572
Conc: 497.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 6.406 min
Delta R.T.: -0.007 min
Response: 334442
Conc: 487.10 ng/ml