

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072021\
 Data File : P0079685.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jul 2021 10:34
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 14:42:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 20 14:31:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.015	4.226	3650733	1239309	49.318	49.326
2) SA Decachlor...	11.140	9.698	2681724	1138853	49.875	50.103
Target Compounds						
26) L6 AR-1254-1	7.262	6.406	1211104	687351	483.326	463.363
27) L6 AR-1254-2	7.492	6.567	1840944	598120	470.927	456.565
28) L6 AR-1254-3	7.876	6.999	1945580	970959	464.653	459.359
29) L6 AR-1254-4	8.173	7.241	1409838	613383	474.983	454.549
30) L6 AR-1254-5	8.603	7.678	1494263	856637	479.723	463.188



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072021\
Data File : P0079685.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jul 2021 10:34
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 14:42:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071921.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 20 14:31:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.015 min
Delta R.T.: 0.000 min
Response: 3650733
Conc: 49.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.226 min
Delta R.T.: -0.012 min
Response: 1239309
Conc: 49.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.140 min
Delta R.T.: -0.002 min
Response: 2681724
Conc: 49.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.698 min
Delta R.T.: -0.006 min
Response: 1138853
Conc: 50.10 ng/ml



#26 AR-1254-1


R.T.: 7.262 min
Delta R.T.: 0.000 min
Response: 1211104
Conc: 483.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.406 min
Delta R.T.: 0.000 min
Response: 687351
Conc: 463.36 ng/ml




#27 AR-1254-2

R.T.: 7.492 min
Delta R.T.: 0.000 min
Response: 1840944
Conc: 470.93 ng/ml



#27 AR-1254-2


R.T.: 6.567 min
Delta R.T.: 0.000 min
Response: 598120
Conc: 456.56 ng/ml



#28 AR-1254-3


R.T.: 7.876 min
Delta R.T.: -0.001 min
Response: 1945580
Conc: 464.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.999 min
Delta R.T.: 0.000 min
Response: 970959
Conc: 459.36 ng/ml




#29 AR-1254-4

R.T.: 8.173 min
Delta R.T.: 0.000 min
Response: 1409838
Conc: 474.98 ng/ml



#29 AR-1254-4


R.T.: 7.241 min
Delta R.T.: 0.000 min
Response: 613383
Conc: 454.55 ng/ml



#30 AR-1254-5

R.T.: 8.603 min
Delta R.T.: -0.001 min
Response: 1494263
Conc: 479.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.678 min
Delta R.T.: 0.000 min
Response: 856637
Conc: 463.19 ng/ml