

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072021\
 Data File : P0079711.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jul 2021 19:08
 Operator : DD\AJ
 Sample : M3070-02 5X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 02-A-02-B-02-C

Manual Integrations
 APPROVED

mohammad
 7/21/2021 3:42:31 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 01:54:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0071921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 20 14:31:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.015	4.233	308259	125251	4.164	4.985
2)	SA Decachlor...	11.142	9.705	239315	109057	4.451	4.798
Target Compounds							
26)	L6 AR-1254-1	7.275	6.418	870132	504417	347.251	340.042
27)	L6 AR-1254-2	7.504	6.572	621887	63265	159.083	48.292 #
28)	L6 AR-1254-3	7.883	7.005	358587	87453	85.639	41.374 #
29)	L6 AR-1254-4	8.174	7.248	105033	39601	35.386	29.346
30)	L6 AR-1254-5	8.604	7.684	75121	58989	24.117	31.895m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072021\
Data File : P0079711.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jul 2021 19:08
Operator : DD\AJ
Sample : M3070-02 5X
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
02-A-02-B-02-C


Manual Integrations
APPROVED

mohammad
7/21/2021 3:42:31 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 01:54:00 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071921.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 20 14:31:26 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 5.015 min
Delta R.T.: 0.000 min
Response: 308259
Conc: 4.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

mohammad
7/21/2021 3:42:31 PM




#1 Tetrachloro-m-xylene

R.T.: 4.233 min
Delta R.T.: -0.005 min
Response: 125251
Conc: 4.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.142 min
Delta R.T.: 0.000 min
Response: 239315
Conc: 4.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.705 min
Delta R.T.: 0.000 min
Response: 109057
Conc: 4.80 ng/ml




#26 AR-1254-1

R.T.: 7.275 min
Delta R.T.: 0.012 min
Response: 870132
Conc: 347.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

mohammad
7/21/2021 3:42:31 PM




#26 AR-1254-1

R.T.: 6.418 min
Delta R.T.: 0.011 min
Response: 504417
Conc: 340.04 ng/ml




#27 AR-1254-2

R.T.: 7.504 min
Delta R.T.: 0.011 min
Response: 621887
Conc: 159.08 ng/ml



#27 AR-1254-2

R.T.: 6.572 min
Delta R.T.: 0.005 min
Response: 63265
Conc: 48.29 ng/ml




#28 AR-1254-3

R.T.: 7.883 min
Delta R.T.: 0.005 min
Response: 358587
Conc: 85.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

mohammad
7/21/2021 3:42:31 PM




#28 AR-1254-3

R.T.: 7.005 min
Delta R.T.: 0.006 min
Response: 87453
Conc: 41.37 ng/ml




#29 AR-1254-4

R.T.: 8.174 min
Delta R.T.: 0.000 min
Response: 105033
Conc: 35.39 ng/ml



#29 AR-1254-4

R.T.: 7.248 min
Delta R.T.: 0.007 min
Response: 39601
Conc: 29.35 ng/ml




#30 AR-1254-5

R.T.: 8.604 min
Delta R.T.: 0.000 min
Response: 75121
Conc: 24.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C

Manual Integrations
APPROVED

mohammad
7/21/2021 3:42:31 PM



#30 AR-1254-5

R.T.: 7.684 min
Delta R.T.: 0.005 min
Response: 58989
Conc: 31.90 ng/ml m