

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072022\
 Data File : P0088192.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2022 10:57
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 12:25:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 15 03:57:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.446	3.605	129.9E6	48675504	50.916	51.591
2) SA Decachlor...	10.299	8.606	107.5E6	52636185	50.605	50.368
Target Compounds						
26) L6 AR-1254-1	6.485	5.485	44630393	30098719	470.719	522.048
27) L6 AR-1254-2	6.705	5.632	67455628	26597315	469.881	518.716
28) L6 AR-1254-3	7.073	6.036	67718312	41123847	462.315	506.765
29) L6 AR-1254-4	7.360	6.265	52801934	25375057	487.878	508.782
30) L6 AR-1254-5	7.782	6.682	57247875	38034862	475.665	503.603


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072022\
Data File : P0088192.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2022 10:57
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 12:25:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 15 03:57:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.446 min
Delta R.T.: 0.000 min
Response: 129896909
Conc: 50.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.605 min
Delta R.T.: 0.000 min
Response: 48675504
Conc: 51.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.299 min
Delta R.T.: 0.000 min
Response: 107470865
Conc: 50.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.606 min
Delta R.T.: -0.003 min
Response: 52636185
Conc: 50.37 ng/ml



#26 AR-1254-1


R.T.: 6.485 min
Delta R.T.: 0.000 min
Response: 44630393
Conc: 470.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.485 min
Delta R.T.: -0.002 min
Response: 30098719
Conc: 522.05 ng/ml




#27 AR-1254-2

R.T.: 6.705 min
Delta R.T.: 0.000 min
Response: 67455628
Conc: 469.88 ng/ml



#27 AR-1254-2


R.T.: 5.632 min
Delta R.T.: -0.002 min
Response: 26597315
Conc: 518.72 ng/ml



#28 AR-1254-3


R.T.: 7.073 min
Delta R.T.: 0.000 min
Response: 67718312
Conc: 462.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.036 min
Delta R.T.: -0.002 min
Response: 41123847
Conc: 506.77 ng/ml




#29 AR-1254-4

R.T.: 7.360 min
Delta R.T.: 0.002 min
Response: 52801934
Conc: 487.88 ng/ml



#29 AR-1254-4


R.T.: 6.265 min
Delta R.T.: -0.002 min
Response: 25375057
Conc: 508.78 ng/ml



#30 AR-1254-5

R.T.: 7.782 min
Delta R.T.: 0.002 min
Response: 57247875
Conc: 475.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.682 min
Delta R.T.: -0.002 min
Response: 38034862
Conc: 503.60 ng/ml