

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0072022\  
 Data File : P0088218.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 21 Jul 2022 02:11  
 Operator : YP/AJ  
 Sample : N3793-10 10X  
 Misc :  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 MOO-2022-28-B

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jul 21 04:37:19 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0071422.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Fri Jul 15 03:57:23 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1   | RT#2  | Resp#1  | Resp#2  | ng/ml   | ng/ml  |
|-----------------------------|--------|-------|---------|---------|---------|--------|
| -----                       |        |       |         |         |         |        |
| System Monitoring Compounds |        |       |         |         |         |        |
| 1) SA Tetrachlo...          | 4.445  | 3.604 | 4987731 | 1891944 | 1.955   | 2.005  |
| 2) SA Decachlor...          | 10.290 | 8.601 | 3894948 | 1930674 | 1.834   | 1.847  |
| Target Compounds            |        |       |         |         |         |        |
| 26) L6 AR-1254-1            | 6.482  | 5.482 | 4046979 | 2889204 | 42.684  | 50.112 |
| 27) L6 AR-1254-2            | 6.702  | 5.630 | 7409418 | 2349197 | 51.612  | 45.815 |
| 28) L6 AR-1254-3            | 7.069  | 6.033 | 5743609 | 3412456 | 39.212m | 42.051 |
| 29) L6 AR-1254-4            | 7.356  | 6.262 | 4915168 | 2249478 | 45.415m | 45.103 |
| 30) L6 AR-1254-5            | 7.777  | 6.679 | 4745303 | 3002142 | 39.428  | 39.750 |
| -----                       |        |       |         |         |         |        |



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0072022\  
Data File : P0088218.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 21 Jul 2022 02:11  
Operator : YP/AJ  
Sample : N3793-10 10X  
Misc :  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
MOO-2022-28-B

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jul 21 04:37:19 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0071422.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Fri Jul 15 03:57:23 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.445 min  
Delta R.T.: -0.002 min  
Response: 4987731  
Conc: 1.96 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
MOO-2022-28-B




#1 Tetrachloro-m-xylene

R.T.: 3.604 min  
Delta R.T.: 0.000 min  
Response: 1891944  
Conc: 2.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.290 min  
Delta R.T.: -0.008 min  
Response: 3894948  
Conc: 1.83 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.601 min  
Delta R.T.: -0.007 min  
Response: 1930674  
Conc: 1.85 ng/ml



#26 AR-1254-1


R.T.: 6.482 min  
Delta R.T.: -0.003 min  
Response: 4046979  
Conc: 42.68 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
MOO-2022-28-B




#26 AR-1254-1

R.T.: 5.482 min  
Delta R.T.: -0.005 min  
Response: 2889204  
Conc: 50.11 ng/ml




#27 AR-1254-2

R.T.: 6.702 min  
Delta R.T.: -0.002 min  
Response: 7409418  
Conc: 51.61 ng/ml



#27 AR-1254-2


R.T.: 5.630 min  
Delta R.T.: -0.005 min  
Response: 2349197  
Conc: 45.82 ng/ml



#28 AR-1254-3


R.T.: 7.069 min  
Delta R.T.: -0.003 min  
Response: 5743609  
Conc: 39.21 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
MOO-2022-28-B




#28 AR-1254-3

R.T.: 6.033 min  
Delta R.T.: -0.005 min  
Response: 3412456  
Conc: 42.05 ng/ml




#29 AR-1254-4

R.T.: 7.356 min  
Delta R.T.: -0.003 min  
Response: 4915168  
Conc: 45.42 ng/ml m



#29 AR-1254-4


R.T.: 6.262 min  
Delta R.T.: -0.004 min  
Response: 2249478  
Conc: 45.10 ng/ml



#30 AR-1254-5

R.T.: 7.777 min  
Delta R.T.: -0.003 min  
Response: 4745303  
Conc: 39.43 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
MOO-2022-28-B



#30 AR-1254-5

R.T.: 6.679 min  
Delta R.T.: -0.005 min  
Response: 3002142  
Conc: 39.75 ng/ml