

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072123\
 Data File : P0096220.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jul 2023 08:41
 Operator : YP/AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO072123AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 20:03:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 22 12:20:47 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.410	3.628	169.6E6	70091630	53.857	53.404
2) SA Decachlor...	10.236	8.663	104.4E6	56627606	53.067	51.978
Target Compounds						
16) L4 AR-1242-1	5.590	4.717	38403254	18885055	509.661	512.198
17) L4 AR-1242-2	5.612	4.736	54906693	25570511	518.145	507.686
18) L4 AR-1242-3	5.676	4.913	34492500	14365039	519.010	512.171
19) L4 AR-1242-4	5.775	4.997	27903127	15237817	512.926	510.073
20) L4 AR-1242-5	6.518	5.522	29360180	17454333	512.645	508.688



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072123\
Data File : P0096220.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jul 2023 08:41
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO072123AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 20:03:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 12:20:47 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.410 min
Delta R.T.: 0.000 min
Response: 169567487
Conc: 53.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO072123AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.628 min
Delta R.T.: -0.001 min
Response: 70091630
Conc: 53.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: -0.001 min
Response: 104423196
Conc: 53.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.663 min
Delta R.T.: 0.000 min
Response: 56627606
Conc: 51.98 ng/ml



#16 AR-1242-1


R.T.: 5.590 min
Delta R.T.: 0.000 min
Response: 38403254
Conc: 509.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO072123AR1242




#16 AR-1242-1

R.T.: 4.717 min
Delta R.T.: 0.000 min
Response: 18885055
Conc: 512.20 ng/ml




#17 AR-1242-2

R.T.: 5.612 min
Delta R.T.: 0.000 min
Response: 54906693
Conc: 518.15 ng/ml



#17 AR-1242-2


R.T.: 4.736 min
Delta R.T.: 0.000 min
Response: 25570511
Conc: 507.69 ng/ml



#18 AR-1242-3


R.T.: 5.676 min
Delta R.T.: 0.000 min
Response: 34492500
Conc: 519.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO072123AR1242




#18 AR-1242-3

R.T.: 4.913 min
Delta R.T.: 0.000 min
Response: 14365039
Conc: 512.17 ng/ml




#19 AR-1242-4

R.T.: 5.775 min
Delta R.T.: 0.000 min
Response: 27903127
Conc: 512.93 ng/ml



#19 AR-1242-4


R.T.: 4.997 min
Delta R.T.: -0.001 min
Response: 15237817
Conc: 510.07 ng/ml



#20 AR-1242-5

R.T.: 6.518 min
Delta R.T.: -0.001 min
Response: 29360180
Conc: 512.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO072123AR1242



#20 AR-1242-5

R.T.: 5.522 min
Delta R.T.: 0.000 min
Response: 17454333
Conc: 508.69 ng/ml