

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072221\
 Data File : P0079764.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jul 2021 17:36
 Operator : DD\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 04:33:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 23 04:31:15 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.002	4.122	1448830	592893	24.983	25.630
2) SA Decachlor...	11.114	9.535	1103797	525700	25.693	26.395
Target Compounds						
26) L6 AR-1254-1	7.247	6.287	504655	337550	257.262	260.962
27) L6 AR-1254-2	7.478	6.448	775883	299933	257.554	263.740
28) L6 AR-1254-3	7.861	6.877	818292	468650	255.291	261.495
29) L6 AR-1254-4	8.159	7.118	580874	294653	254.424	257.675
30) L6 AR-1254-5	8.588	7.553	631814	387762	259.113	260.504



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072221\
Data File : P0079764.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Jul 2021 17:36
Operator : DD\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 04:33:14 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 23 04:31:15 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 1448830
Conc: 24.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 4.122 min
Delta R.T.: 0.000 min
Response: 592893
Conc: 25.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.114 min
Delta R.T.: 0.000 min
Response: 1103797
Conc: 25.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.535 min
Delta R.T.: 0.000 min
Response: 525700
Conc: 26.39 ng/ml



#26 AR-1254-1


R.T.: 7.247 min
Delta R.T.: 0.000 min
Response: 504655
Conc: 257.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 6.287 min
Delta R.T.: 0.000 min
Response: 337550
Conc: 260.96 ng/ml




#27 AR-1254-2

R.T.: 7.478 min
Delta R.T.: 0.000 min
Response: 775883
Conc: 257.55 ng/ml



#27 AR-1254-2


R.T.: 6.448 min
Delta R.T.: 0.000 min
Response: 299933
Conc: 263.74 ng/ml



#28 AR-1254-3


R.T.: 7.861 min
Delta R.T.: 0.000 min
Response: 818292
Conc: 255.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.877 min
Delta R.T.: 0.000 min
Response: 468650
Conc: 261.50 ng/ml




#29 AR-1254-4

R.T.: 8.159 min
Delta R.T.: 0.000 min
Response: 580874
Conc: 254.42 ng/ml



#29 AR-1254-4


R.T.: 7.118 min
Delta R.T.: 0.000 min
Response: 294653
Conc: 257.67 ng/ml



#30 AR-1254-5

R.T.: 8.588 min
Delta R.T.: 0.000 min
Response: 631814
Conc: 259.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 7.553 min
Delta R.T.: 0.000 min
Response: 387762
Conc: 260.50 ng/ml