

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072224\
 Data File : P0104896.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jul 2024 16:03
 Operator : YP/AJ
 Sample : PB162192BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB162192BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/23/2024
 Supervised By :Ankita Jodhani 07/23/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 01:25:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 02 16:29:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.427	3.541	173.1E6	63184852	19.334	19.366
2) SA Decachlor...	10.212	8.435	123.7E6	38669629	20.166m	18.440

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072224\
Data File : PO104896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jul 2024 16:03
Operator : YP/AJ
Sample : PB162192BL
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PB162192BL


Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 07/23/2024
Supervised By : Ankita Jodhani 07/23/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 01:25:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO070224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 02 16:29:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.427 min
Delta R.T.: 0.002 min
Response: 173102388
Conc: 19.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB162192BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/23/2024
Supervised By :Ankita Jodhani 07/23/2024




#1 Tetrachloro-m-xylene

R.T.: 3.541 min
Delta R.T.: -0.001 min
Response: 63184852
Conc: 19.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.212 min
Delta R.T.: 0.006 min
Response: 123668748
Conc: 20.17 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.435 min
Delta R.T.: -0.006 min
Response: 38669629
Conc: 18.44 ng/ml