

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\methods\
 Method File : P0072318.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jul 24 03:04:21 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0047008.D 1000 =P0047006.D 750 =P0047007.D
 250 =P0047009.D 50 =P0047010.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		1.708	1.698	1.705	1.795	1.477	1.677	E5	7.06
2) SA	Decachlorobiphenyl		1.145	1.108	1.125	1.168	0.927	1.095	E5	8.79
3) L1	AR-1016-1		3.363	3.240	3.291	3.498	2.871	3.253	E3	7.20
4) L1	AR-1016-2		4.002	3.906	3.958	4.153	3.540	3.912	E3	5.81
5) L1	AR-1016-3		3.253	3.190	3.211	3.379	3.066	3.219	E3	3.51
6) L1	AR-1016-4		3.261	3.144	3.186	3.333	2.852	3.155	E3	5.83
7) L1	AR-1016-5		3.430	3.313	3.343	3.455	2.972	3.303	E3	5.87
8) L2	AR-1221-1		1.623					1.623	E3	0.00
9) L2	AR-1221-2		1.188					1.188	E3	0.00
10) L2	AR-1221-3		3.707					3.707	E3	0.00
11) L3	AR-1232-1		1.498					1.498	E3	0.00
12) L3	AR-1232-2		2.038					2.038	E3	0.00
13) L3	AR-1232-3		2.908					2.908	E3	0.00
14) L3	AR-1232-4		1.784					1.784	E3	0.00
15) L3	AR-1232-5		1.417					1.417	E3	0.00
16) L4	AR-1242-1		5.796					5.796	E3	0.00
17) L4	AR-1242-2		3.547					3.547	E3	0.00
18) L4	AR-1242-3		2.838					2.838	E3	0.00
19) L4	AR-1242-4		2.876					2.876	E3	0.00
20) L4	AR-1242-5		3.098					3.098	E3	0.00
21) L5	AR-1248-1		4.676					4.676	E3	0.00
22) L5	AR-1248-2		3.963					3.963	E3	0.00
23) L5	AR-1248-3		5.077					5.077	E3	0.00
24) L5	AR-1248-4		4.777					4.777	E3	0.00
25) L5	AR-1248-5		3.203					3.203	E3	0.00
26) L6	AR-1254-1		8.222					8.222	E3	0.00
27) L6	AR-1254-2		4.910					4.910	E3	0.00
28) L6	AR-1254-3		8.594					8.594	E3	0.00
29) L6	AR-1254-4		6.387					6.387	E3	0.00
30) L6	AR-1254-5		4.748					4.748	E3	0.00
31) L7	AR-1260-1		6.181	5.967	6.018	6.342	5.519	6.005	E3	5.15
32) L7	AR-1260-2		7.480	7.241	7.308	7.643	6.866	7.308	E3	4.00
33) L7	AR-1260-3		8.733	8.552	8.592	8.906	8.091	8.575	E3	3.54
34) L7	AR-1260-4		1.251	1.225	1.226	1.251	1.100	1.211	E4	5.22
35) L7	AR-1260-5		7.623	7.473	7.496	7.652	6.641	7.377	E3	5.68
36) L8	AR-1262-1		5.382					5.382	E3	0.00
37) L8	AR-1262-2		6.437					6.437	E3	0.00
38) L8	AR-1262-3		8.252					8.252	E3	0.00
39) L8	AR-1262-4		7.989					7.989	E3	0.00
40) L8	AR-1262-5		1.490					1.490	E4	0.00
41) L9	AR-1268-1		1.566					1.566	E4	0.00
42) L9	AR-1268-2		1.401					1.401	E4	0.00
43) L9	AR-1268-3		1.166					1.166	E4	0.00
44) L9	AR-1268-4		5.053					5.053	E3	0.00
45) L9	AR-1268-5		3.659					3.659	E4	0.00

Signal #2 Calibration Files

500 =P0047008.D 1000 =P0047006.D 750 =P0047007.D
 250 =P0047009.D 50 =P0047010.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		1.934	1.878	1.921	2.083	1.774	1.918	E5	5.81
2) SA	Decachlorobiphenyl		9.492	9.064	9.216	9.613	8.536	9.184	E4	4.60

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\methods\
 Method File : P0072318.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jul 24 03:04:21 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0047008.D 1000 =P0047006.D 750 =P0047007.D
 250 =P0047009.D 50 =P0047010.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	3.508	3.294	3.429	3.789	2.967	3.397	E3	8.85
4)	L1	AR-1016-2	3.365	3.164	3.279	3.413	2.864	3.217	E3	6.80
5)	L1	AR-1016-3	2.689	2.491	2.639	2.844	2.464	2.625	E3	5.90
6)	L1	AR-1016-4	3.251	3.054	3.168	3.360	2.721	3.111	E3	7.87
7)	L1	AR-1016-5	3.451	3.197	3.365	3.690	3.563	3.453	E3	5.46
8)	L2	AR-1221-1	1.666					1.666	E3	0.00
9)	L2	AR-1221-2	1.198					1.198	E3	0.00
10)	L2	AR-1221-3	3.948					3.948	E3	0.00
11)	L3	AR-1232-1	1.572					1.572	E3	0.00
12)	L3	AR-1232-2	1.994					1.994	E3	0.00
13)	L3	AR-1232-3	3.091					3.091	E3	0.00
14)	L3	AR-1232-4	1.823					1.823	E3	0.00
15)	L3	AR-1232-5	1.005					1.005	E3	0.00
16)	L4	AR-1242-1	5.753					5.753	E3	0.00
17)	L4	AR-1242-2	2.877					2.877	E3	0.00
18)	L4	AR-1242-3	2.857					2.857	E3	0.00
19)	L4	AR-1242-4	3.055					3.055	E3	0.00
20)	L4	AR-1242-5	3.231					3.231	E3	0.00
21)	L5	AR-1248-1	5.013					5.013	E3	0.00
22)	L5	AR-1248-2	4.191					4.191	E3	0.00
23)	L5	AR-1248-3	6.411					6.411	E3	0.00
24)	L5	AR-1248-4	4.476					4.476	E3	0.00
25)	L5	AR-1248-5	2.767					2.767	E3	0.00
26)	L6	AR-1254-1	7.144					7.144	E3	0.00
27)	L6	AR-1254-2	6.057					6.057	E3	0.00
28)	L6	AR-1254-3	9.415					9.415	E3	0.00
29)	L6	AR-1254-4	5.768					5.768	E3	0.00
30)	L6	AR-1254-5	3.970					3.970	E3	0.00
31)	L7	AR-1260-1	5.959	5.547	5.766	6.252	6.115	5.928	E3	4.71
32)	L7	AR-1260-2	7.080	6.615	6.829	7.375	7.219	7.023	E3	4.33
33)	L7	AR-1260-3	7.724	7.293	7.457	7.894	7.170	7.508	E3	3.99
34)	L7	AR-1260-4	1.193	1.139	1.162	1.208	1.128	1.166	E4	2.94
35)	L7	AR-1260-5	8.387	8.071	8.154	8.424	8.425	8.292	E3	2.02
36)	L8	AR-1262-1	6.171					6.171	E3	0.00
37)	L8	AR-1262-2	6.121					6.121	E3	0.00
38)	L8	AR-1262-3	8.139					8.139	E3	0.00
39)	L8	AR-1262-4	7.098					7.098	E3	0.00
40)	L8	AR-1262-5	1.400					1.400	E4	0.00
41)	L9	AR-1268-1	1.363					1.363	E4	0.00
42)	L9	AR-1268-2	1.255					1.255	E4	0.00
43)	L9	AR-1268-3	1.006					1.006	E4	0.00
44)	L9	AR-1268-4	4.380					4.380	E3	0.00
45)	L9	AR-1268-5	3.002					3.002	E4	0.00

(#) = Out of Range