

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072320\
 Data File : P0069882.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jul 2020 16:05
 Operator : DD\AJ
 Sample : P0072320ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 02:20:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 23 02:12:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	34.327	35.599 E3	-3.7	103	0.00
2 SA Decachlorobiphenyl	53.432	54.327 E3	-1.7	102	0.00
3 L1 AR-1016-1	1.562	1.607 E3	-2.9	104	0.00
4 L1 AR-1016-2	2.202	2.311 E3	-5.0	104	0.00
5 L1 AR-1016-3	1.330	1.380 E3	-3.8	105	0.00
6 L1 AR-1016-4	1.155	1.175 E3	-1.7	105	0.00
7 L1 AR-1016-5	1.124	1.155 E3	-2.8	104	0.00
31 L7 AR-1260-1	2.154	2.302 E3	-6.9	106	0.00
32 L7 AR-1260-2	3.008	3.221 E3	-7.1	107	0.00
33 L7 AR-1260-3	2.609	2.734 E3	-4.8	105	0.00
34 L7 AR-1260-4	2.505	2.599 E3	-3.8	105	0.00
35 L7 AR-1260-5	5.848	6.260 E3	-7.0	106	0.00

Signal #2

1 SA Tetrachloro-m-xylene	36.044	37.070 E3	-2.8	103	0.00
2 SA Decachlorobiphenyl	51.002	50.854 E3	0.3	104	0.00
3 L1 AR-1016-1	1.585	1.594 E3	-0.6	103	0.00
4 L1 AR-1016-2	2.211	2.211 E3	0.0	102	0.00
5 L1 AR-1016-3	1.187	1.198 E3	-0.9	103	0.00
6 L1 AR-1016-4	1005.869	1020.810	-1.5	103	0.00
7 L1 AR-1016-5	1.228	1.250 E3	-1.8	102	0.00
31 L7 AR-1260-1	2.414	2.354 E3	2.5	101	0.00
32 L7 AR-1260-2	3.015	2.991 E3	0.8	102	0.00
33 L7 AR-1260-3	2.754	2.715 E3	1.4	102	0.00
34 L7 AR-1260-4	2.387	2.374 E3	0.5	102	0.00
35 L7 AR-1260-5	5.931	5.987 E3	-0.9	103	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	392.563	0.000	100.0#	0#	-4.64#
9 L2 AR-1221-2	303.066	0.000	100.0#	0#	-4.73#
10 L2 AR-1221-3	1.007	0.000 E3	100.0#	0#	-4.82#
11 L3 AR-1232-1	833.453	0.000	100.0#	0#	-4.82#
12 L3 AR-1232-2	431.000	0.000	100.0#	0#	-5.39#
13 L3 AR-1232-3	983.436	0.000	100.0#	0#	-5.71#
14 L3 AR-1232-4	494.672	0.000	100.0#	0#	-5.88#
15 L3 AR-1232-5	320.022	0.000	100.0#	0#	-5.98#
16 L4 AR-1242-1	1.243	0.000 E3	100.0#	0#	-5.69#
17 L4 AR-1242-2	1.786	0.000 E3	100.0#	0#	-5.71#
18 L4 AR-1242-3	1.083	0.000 E3	100.0#	0#	-5.77#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072320\
 Data File : P0069882.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jul 2020 16:05
 Operator : DD\AJ
 Sample : P0072320ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 02:20:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 23 02:12:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	913.162	0.000	100.0#	0#	-5.88#
20 L4 AR-1242-5	1.149	0.000 E3	100.0#	0#	-6.65#
21 L5 AR-1248-1	868.867	0.000	100.0#	0#	-5.69#
22 L5 AR-1248-2	1.124	0.000 E3	100.0#	0#	-5.98#
23 L5 AR-1248-3	1.363	0.000 E3	100.0#	0#	-6.19#
24 L5 AR-1248-4	1.753	0.000 E3	100.0#	0#	-6.62#
25 L5 AR-1248-5	1.670	0.000 E3	100.0#	0#	-6.65#
26 L6 AR-1254-1	1.765	0.000 E3	100.0#	0#	-6.58#
27 L6 AR-1254-2	2.811	0.000 E3	100.0#	0#	-6.81#
28 L6 AR-1254-3	3.037	0.000 E3	100.0#	0#	-7.19#
29 L6 AR-1254-4	2.736	0.000 E3	100.0#	0#	-7.48#
30 L6 AR-1254-5	2.640	0.000 E3	100.0#	0#	-7.91#
36 L8 AR-1262-1	3.715	0.000 E3	100.0#	0#	-7.98#
37 L8 AR-1262-2	6.610	0.000 E3	100.0#	0#	-8.52#
38 L8 AR-1262-3	2.981	0.000 E3	100.0#	0#	-8.79#
39 L8 AR-1262-4	3.256	0.000 E3	100.0#	0#	-8.87#
40 L8 AR-1262-5	2.404	0.000 E3	100.0#	0#	-9.43#
41 L9 AR-1268-1	8.011	0.000 E3	100.0#	0#	-8.79#
42 L9 AR-1268-2	7.049	0.000 E3	100.0#	0#	-8.87#
43 L9 AR-1268-3	6.121	0.000 E3	100.0#	0#	-9.05#
44 L9 AR-1268-4	2.663	0.000 E3	100.0#	0#	-9.43#
45 L9 AR-1268-5	19.221	0.000 E3	100.0#	0#	-9.76#

Signal #2

8 L2 AR-1221-1	445.756	0.000	100.0#	0#	-3.81#
9 L2 AR-1221-2	333.299	0.000	100.0#	0#	-3.90#
10 L2 AR-1221-3	1.041	0.000 E3	100.0#	0#	-3.99#
11 L3 AR-1232-1	870.095	0.000	100.0#	0#	-3.99#
12 L3 AR-1232-2	972.586	0.000	100.0#	0#	-4.81#
13 L3 AR-1232-3	517.929	0.000	100.0#	0#	-4.99#
14 L3 AR-1232-4	425.954	0.000	100.0#	0#	-5.09#
15 L3 AR-1232-5	506.912	0.000	100.0#	0#	-5.27#
16 L4 AR-1242-1	1.239	0.000 E3	100.0#	0#	-4.79#
17 L4 AR-1242-2	1.717	0.000 E3	100.0#	0#	-4.81#
18 L4 AR-1242-3	929.408	0.000	100.0#	0#	-4.99#
19 L4 AR-1242-4	843.959	0.000	100.0#	0#	-5.09#
20 L4 AR-1242-5	1.229	0.000 E3	100.0#	0#	-5.65#
21 L5 AR-1248-1	947.499	0.000	100.0#	0#	-4.79#
22 L5 AR-1248-2	1.294	0.000 E3	100.0#	0#	-5.05#
23 L5 AR-1248-3	1.193	0.000 E3	100.0#	0#	-5.09#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072320\
 Data File : P0069882.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jul 2020 16:05
 Operator : DD\AJ
 Sample : P0072320ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 02:20:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 23 02:12:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	1.544	0.000 E3	100.0#	0#	-5.27#
25 L5 AR-1248-5	1.579	0.000 E3	100.0#	0#	-5.69#
26 L6 AR-1254-1	2.550	0.000 E3	100.0#	0#	-5.65#
27 L6 AR-1254-2	2.255	0.000 E3	100.0#	0#	-5.81#
28 L6 AR-1254-3	3.584	0.000 E3	100.0#	0#	-6.22#
29 L6 AR-1254-4	2.516	0.000 E3	100.0#	0#	-6.46#
30 L6 AR-1254-5	3.417	0.000 E3	100.0#	0#	-6.88#
36 L8 AR-1262-1	1.783	0.000 E3	100.0#	0#	-6.88#
37 L8 AR-1262-2	6.252	0.000 E3	100.0#	0#	-7.43#
38 L8 AR-1262-3	2.615	0.000 E3	100.0#	0#	-7.72#
39 L8 AR-1262-4	4.744	0.000 E3	100.0#	0#	-7.78#
40 L8 AR-1262-5	2.366	0.000 E3	100.0#	0#	-8.27#
41 L9 AR-1268-1	7.265	0.000 E3	100.0#	0#	-7.72#
42 L9 AR-1268-2	6.698	0.000 E3	100.0#	0#	-7.78#
43 L9 AR-1268-3	5.596	0.000 E3	100.0#	0#	-7.98#
44 L9 AR-1268-4	2.626	0.000 E3	100.0#	0#	-8.27#
45 L9 AR-1268-5	17.901	0.000 E3	100.0#	0#	-8.54#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0