

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072320\
 Data File : P0069898.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jul 2020 20:40
 Operator : DD\AJ
 Sample : L3216-07
 Misc : AR1232 LOD 25 PPB
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 02:25:51 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 23 02:12:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds						
1) SA Tetrachlo...	4.381	3.556	535073	555155	15.587	15.402
2) SA Decachlor...	10.029	8.759	852528	913059	15.956	17.903
Target Compounds						
43) L9 AR-1268-3	9.082f	0.000	442946	0	72.367	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0072320\
Data File : P0069898.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Jul 2020 20:40
Operator : DD\AJ
Sample : L3216-07
Misc : AR1232 LOD 25 PPB
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 02:25:51 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072320.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 23 02:12:36 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped


Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.381 min
Delta R.T.: 0.000 min
Response: 535073
Conc: 15.59 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 3.556 min
Delta R.T.: 0.000 min
Response: 555155
Conc: 15.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.029 min
Delta R.T.: -0.002 min
Response: 852528
Conc: 15.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.759 min
Delta R.T.: 0.000 min
Response: 913059
Conc: 17.90 ng/ml



#43 AR-1268-3

R.T.: 9.082 min
Delta R.T.: 0.028 min
Response: 442946
Conc: 72.37 ng/ml



#43 AR-1268-3

R.T.: 0.000 min
Exp R.T.: 7.982 min
Response: 0
Conc: N.D.