

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072320\
 Data File : P0069871.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jul 2020 12:39
 Operator : DD\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 00:51:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 23 00:49:41 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.381	3.556	2483198	2569178	71.839	71.283
2)	SA Decachlor...	10.029	8.759	3900160	3484655	72.875	71.243
Target Compounds							
3)	L1 AR-1016-1	5.688	4.788	1091700	1074025	707.972	692.615
4)	L1 AR-1016-2	5.711	4.806	1559763	1517647	703.789	702.498
5)	L1 AR-1016-3	5.773	4.992	928651	812239	704.437	696.691
6)	L1 AR-1016-4	5.879	5.048	787390	694116	705.551	697.179
7)	L1 AR-1016-5	6.190	5.270	772784	854641	698.226	698.744
31)	L7 AR-1260-1	7.351	6.354	1524392	1624955	704.435	697.898
32)	L7 AR-1260-2	7.618	6.555	2160785	2051391	716.418	702.777
33)	L7 AR-1260-3	7.977	6.703	1865147	1878716	712.832	704.729
34)	L7 AR-1260-4	8.203	7.181	1771289	1629597	718.505	701.484
35)	L7 AR-1260-5	8.517	7.432	4298443	4170217	726.585	718.461



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072320\
Data File : P0069871.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Jul 2020 12:39
Operator : DD\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 00:51:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072320.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 23 00:49:41 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.381 min
Delta R.T.: 0.000 min
Response: 2483198
Conc: 71.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.556 min
Delta R.T.: 0.000 min
Response: 2569178
Conc: 71.28 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.029 min
Delta R.T.: -0.001 min
Response: 3900160
Conc: 72.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.759 min
Delta R.T.: 0.000 min
Response: 3484655
Conc: 71.24 ng/ml



#3 AR-1016-1


R.T.: 5.688 min
Delta R.T.: 0.000 min
Response: 1091700
Conc: 707.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 4.788 min
Delta R.T.: 0.000 min
Response: 1074025
Conc: 692.62 ng/ml




#4 AR-1016-2

R.T.: 5.711 min
Delta R.T.: 0.001 min
Response: 1559763
Conc: 703.79 ng/ml



#4 AR-1016-2


R.T.: 4.806 min
Delta R.T.: 0.000 min
Response: 1517647
Conc: 702.50 ng/ml



#5 AR-1016-3


R.T.: 5.773 min
Delta R.T.: 0.000 min
Response: 928651
Conc: 704.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 4.992 min
Delta R.T.: 0.000 min
Response: 812239
Conc: 696.69 ng/ml




#6 AR-1016-4

R.T.: 5.879 min
Delta R.T.: 0.000 min
Response: 787390
Conc: 705.55 ng/ml



#6 AR-1016-4


R.T.: 5.048 min
Delta R.T.: 0.000 min
Response: 694116
Conc: 697.18 ng/ml



#7 AR-1016-5


R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 772784
Conc: 698.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.270 min
Delta R.T.: 0.000 min
Response: 854641
Conc: 698.74 ng/ml




#31 AR-1260-1

R.T.: 7.351 min
Delta R.T.: 0.000 min
Response: 1524392
Conc: 704.43 ng/ml



#31 AR-1260-1


R.T.: 6.354 min
Delta R.T.: 0.000 min
Response: 1624955
Conc: 697.90 ng/ml



#32 AR-1260-2


R.T.: 7.618 min
Delta R.T.: 0.000 min
Response: 2160785
Conc: 716.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.555 min
Delta R.T.: 0.000 min
Response: 2051391
Conc: 702.78 ng/ml




#33 AR-1260-3

R.T.: 7.977 min
Delta R.T.: 0.000 min
Response: 1865147
Conc: 712.83 ng/ml



#33 AR-1260-3


R.T.: 6.703 min
Delta R.T.: 0.000 min
Response: 1878716
Conc: 704.73 ng/ml



#34 AR-1260-4


R.T.: 8.203 min
Delta R.T.: 0.000 min
Response: 1771289
Conc: 718.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 7.181 min
Delta R.T.: 0.000 min
Response: 1629597
Conc: 701.48 ng/ml



#35 AR-1260-5

R.T.: 8.517 min
Delta R.T.: 0.000 min
Response: 4298443
Conc: 726.58 ng/ml



#35 AR-1260-5

R.T.: 7.432 min
Delta R.T.: 0.000 min
Response: 4170217
Conc: 718.46 ng/ml