

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072321\
 Data File : P0079809.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Jul 2021 19:47
 Operator : DD\AJ
 Sample : M2940-01
 Misc : AR1262 LOD 25 PPB
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOD-MDL-SOIL-01-QT3-2021

Manual Integrations
 APPROVED

mohammad
 7/26/2021 11:57:57 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 04:29:21 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 23 05:30:42 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.002	4.121	893389	374049	15.799	16.824
2) SA Decachlor...	11.112	9.533	729904	358927	18.470	20.927
Target Compounds						
36) L8 AR-1262-1	8.659	7.549	68193	26116	23.582m	29.595 #
37) L8 AR-1262-2	9.227	8.108	127813	70361	23.852	25.504
38) L8 AR-1262-3	9.545	8.399	82197	34108	22.601m	30.381 #
39) L8 AR-1262-4	9.622f	8.463	70822	55238	25.048	27.313
40) L8 AR-1262-5	10.334	8.970	42403	20695	22.589m	21.972

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072321\
Data File : P0079809.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Jul 2021 19:47
Operator : DD\AJ
Sample : M2940-01
Misc : AR1262 LOD 25 PPB
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
LOD-MDL-SOIL-01-QT3-2021

Manual Integrations
APPROVED

mohammad
7/26/2021 11:57:57 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 26 04:29:21 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 23 05:30:42 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

