

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072420\
 Data File : P0069963.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Jul 2020 2:01
 Operator : DD\AJ
 Sample : L3216-07
 Misc : AR1660 LOD 40 PPB
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2020

Manual Integrations
 APPROVED

Ankita
 7/24/2020 2:45:00 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 04:56:47 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 23 02:12:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.376	3.554	699990	742723	20.392	20.606
2) SA Decachlor...	10.017	8.755	1215063	1144305	22.741	22.437
Target Compounds						
3) L1 AR-1016-1	5.682	4.785	70061	79525	44.865	50.177m
4) L1 AR-1016-2	5.704	4.804	104724	104357	47.557	47.200
5) L1 AR-1016-3	5.767	4.989	61293	60127	46.068	50.666
6) L1 AR-1016-4	5.873	5.047	52129	50172	45.117	49.880
7) L1 AR-1016-5	6.184	5.267	48651	63410	43.285	51.634
31) L7 AR-1260-1	7.342	6.351	114268	132930	53.046	55.077
32) L7 AR-1260-2	7.609	6.552	163679	164898	54.414	54.697
33) L7 AR-1260-3	7.969	6.700	139733	145820	53.567	52.940
34) L7 AR-1260-4	8.194	7.177	132701	118559	52.977	49.666
35) L7 AR-1260-5	8.509	7.429	303015	304891	51.812	51.404

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072420\
Data File : P0069963.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Jul 2020 2:01
Operator : DD\AJ
Sample : L3216-07
Misc : AR1660 LOD 40 PPB
ALS Vial : 48 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT3-2020

Manual Integrations
APPROVED

Ankita
7/24/2020 2:45:00 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 04:56:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072320.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 23 02:12:36 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

