

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072423\
 Data File : P0096280.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2023 15:34
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 21:45:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 22 12:20:47 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.411	3.627	159.0E6	64077764	50.498	48.822
2)	SA Decachlor...	10.235	8.658	109.5E6	55107578	55.669	50.583
Target Compounds							
26)	L6 AR-1254-1	6.454	5.521	50417361	33109775	504.562	481.111
27)	L6 AR-1254-2	6.675	5.670	73739100	29170002	509.591	480.376
28)	L6 AR-1254-3	7.044	6.075	70890109	45487060	492.336	487.838
29)	L6 AR-1254-4	7.331	6.305	47226949	27115762	509.311	485.341
30)	L6 AR-1254-5	7.753	6.724	53718121	39384739	522.595	494.534

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072423\
Data File : P0096280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2023 15:34
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 21:45:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 12:20:47 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

