

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072423\
Data File : P0096278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2023 14:44
Operator : YP/AJ
Sample : 03719-06
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
B-P23B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 21:43:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 12:20:47 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.411	3.627	67189545	27664616	21.340	21.078
2) SA Decachlor...	10.236	8.659	46428936	25308966	23.595	23.231

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072423\
Data File : P0096278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2023 14:44
Operator : YP/AJ
Sample : 03719-06
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
B-P23B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 21:43:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 12:20:47 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.411 min
Delta R.T.: 0.001 min
Response: 67189545
Conc: 21.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
B-P23B




#1 Tetrachloro-m-xylene

R.T.: 3.627 min
Delta R.T.: -0.002 min
Response: 27664616
Conc: 21.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 46428936
Conc: 23.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.659 min
Delta R.T.: -0.005 min
Response: 25308966
Conc: 23.23 ng/ml