

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072424\
 Data File : P0104925.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2024 10:26
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 12:40:02 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 02 16:29:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.428	3.539	424.4E6	155.3E6	47.401	47.586
2) SA Decachlor...	10.215	8.434	273.9E6	84593312	44.661	40.338
Target Compounds						
3) L1 AR-1016-1	5.598	4.595	129.6E6	53091827	435.553	468.134
4) L1 AR-1016-2	5.620	4.613	196.3E6	72049172	471.622	463.715
5) L1 AR-1016-3	5.683	4.785	119.4E6	39666951	468.182	480.548
6) L1 AR-1016-4	5.782	4.827	102.1E6	31714068	476.543	487.880
7) L1 AR-1016-5	6.079	5.036	94011494	40972570	450.128	463.556
31) L7 AR-1260-1	7.212	6.051	170.9E6	74336888	473.445	464.231
32) L7 AR-1260-2	7.470	6.237	210.5E6	88602488	498.955	461.209
33) L7 AR-1260-3	7.833	6.387	142.8E6	81717117	483.245	459.445
34) L7 AR-1260-4	8.060	6.853	187.6E6	59944085	529.556	447.451
35) L7 AR-1260-5	8.383	7.092	334.6E6	135.0E6	513.336	441.201



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072424\
Data File : PO104925.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2024 10:26
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 12:40:02 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO070224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 02 16:29:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.428 min
Delta R.T.: 0.003 min
Response: 424399228
Conc: 47.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.539 min
Delta R.T.: -0.003 min
Response: 155252703
Conc: 47.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.215 min
Delta R.T.: 0.009 min
Response: 273884231
Conc: 44.66 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.434 min
Delta R.T.: -0.006 min
Response: 84593312
Conc: 40.34 ng/ml



#3 AR-1016-1


R.T.: 5.598 min
Delta R.T.: 0.006 min
Response: 129588488
Conc: 435.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.595 min
Delta R.T.: -0.004 min
Response: 53091827
Conc: 468.13 ng/ml




#4 AR-1016-2

R.T.: 5.620 min
Delta R.T.: 0.005 min
Response: 196334397
Conc: 471.62 ng/ml



#4 AR-1016-2


R.T.: 4.613 min
Delta R.T.: -0.004 min
Response: 72049172
Conc: 463.72 ng/ml



#5 AR-1016-3


R.T.: 5.683 min
Delta R.T.: 0.005 min
Response: 119406541
Conc: 468.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.785 min
Delta R.T.: -0.004 min
Response: 39666951
Conc: 480.55 ng/ml




#6 AR-1016-4

R.T.: 5.782 min
Delta R.T.: 0.006 min
Response: 102134096
Conc: 476.54 ng/ml



#6 AR-1016-4


R.T.: 4.827 min
Delta R.T.: -0.004 min
Response: 31714068
Conc: 487.88 ng/ml



#7 AR-1016-5


R.T.: 6.079 min
Delta R.T.: 0.006 min
Response: 94011494
Conc: 450.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.036 min
Delta R.T.: -0.004 min
Response: 40972570
Conc: 463.56 ng/ml




#31 AR-1260-1

R.T.: 7.212 min
Delta R.T.: 0.007 min
Response: 170914895
Conc: 473.45 ng/ml



#31 AR-1260-1


R.T.: 6.051 min
Delta R.T.: -0.004 min
Response: 74336888
Conc: 464.23 ng/ml



#32 AR-1260-2


R.T.: 7.470 min
Delta R.T.: 0.007 min
Response: 210512736
Conc: 498.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.237 min
Delta R.T.: -0.004 min
Response: 88602488
Conc: 461.21 ng/ml




#33 AR-1260-3

R.T.: 7.833 min
Delta R.T.: 0.007 min
Response: 142769651
Conc: 483.25 ng/ml



#33 AR-1260-3


R.T.: 6.387 min
Delta R.T.: -0.004 min
Response: 81717117
Conc: 459.45 ng/ml



#34 AR-1260-4


R.T.: 8.060 min
Delta R.T.: 0.006 min
Response: 187618398
Conc: 529.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.853 min
Delta R.T.: -0.005 min
Response: 59944085
Conc: 447.45 ng/ml



#35 AR-1260-5

R.T.: 8.383 min
Delta R.T.: 0.007 min
Response: 334571697
Conc: 513.34 ng/ml



#35 AR-1260-5

R.T.: 7.092 min
Delta R.T.: -0.005 min
Response: 135031890
Conc: 441.20 ng/ml