

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072425\
 Data File : P0112451.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2025 09:39
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 23:32:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.670	3.662	449.2E6	266.2E6	55.238	53.565
2) SA Decachlor...	8.694	8.640	396.6E6	90846120	54.242	51.409
Target Compounds						
21) L5 AR-1248-1	4.756	4.737	96337339	59464958	514.599	489.386
22) L5 AR-1248-2	4.994	4.973	129.5E6	83663876	507.491	494.235
23) L5 AR-1248-3	5.208	5.014	174.5E6	88846459	507.686	500.129
24) L5 AR-1248-4	5.561	5.185	261.6E6	106.0E6	514.799	508.751
25) L5 AR-1248-5	5.602	5.575	177.3E6	107.2E6	507.152	498.924



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072425\
Data File : P0112451.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2025 09:39
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 23:32:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.670 min
Delta R.T.: 0.000 min
Response: 449168595
Conc: 55.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.662 min
Delta R.T.: 0.000 min
Response: 266172990
Conc: 53.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.694 min
Delta R.T.: 0.001 min
Response: 396609252
Conc: 54.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.640 min
Delta R.T.: 0.000 min
Response: 90846120
Conc: 51.41 ng/ml



#21 AR-1248-1


R.T.: 4.756 min
Delta R.T.: 0.000 min
Response: 96337339
Conc: 514.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.737 min
Delta R.T.: 0.000 min
Response: 59464958
Conc: 489.39 ng/ml




#22 AR-1248-2

R.T.: 4.994 min
Delta R.T.: 0.000 min
Response: 129467305
Conc: 507.49 ng/ml



#22 AR-1248-2


R.T.: 4.973 min
Delta R.T.: 0.000 min
Response: 83663876
Conc: 494.23 ng/ml



#23 AR-1248-3


R.T.: 5.208 min
Delta R.T.: 0.000 min
Response: 174492439
Conc: 507.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.014 min
Delta R.T.: 0.000 min
Response: 88846459
Conc: 500.13 ng/ml




#24 AR-1248-4

R.T.: 5.561 min
Delta R.T.: 0.000 min
Response: 261558012
Conc: 514.80 ng/ml



#24 AR-1248-4


R.T.: 5.185 min
Delta R.T.: 0.000 min
Response: 106000679
Conc: 508.75 ng/ml



#25 AR-1248-5

R.T.: 5.602 min
Delta R.T.: 0.000 min
Response: 177295841
Conc: 507.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.575 min
Delta R.T.: -0.001 min
Response: 107151530
Conc: 498.92 ng/ml