

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072522\
 Data File : P0088299.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jul 2022 13:15
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 03:29:59 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 15 03:57:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.447	3.605	137.6E6	51190372	53.946	54.257
2) SA Decachlor...	10.295	8.602	96728575	48679540	45.547	46.582
Target Compounds						
26) L6 AR-1254-1	6.484	5.482	45270142	31036939	477.466	538.321
27) L6 AR-1254-2	6.703	5.630	69473083	27183704	483.934	530.152
28) L6 AR-1254-3	7.071	6.033	71457652	41354463	487.843	509.607
29) L6 AR-1254-4	7.359	6.262	52436255	25106475	484.499	503.397
30) L6 AR-1254-5	7.779	6.680	53981924	37244480	448.529	493.138



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072522\
Data File : P0088299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jul 2022 13:15
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 26 03:29:59 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 15 03:57:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.447 min
Delta R.T.: 0.000 min
Response: 137627349
Conc: 53.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.605 min
Delta R.T.: 0.000 min
Response: 51190372
Conc: 54.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.295 min
Delta R.T.: -0.003 min
Response: 96728575
Conc: 45.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.602 min
Delta R.T.: -0.006 min
Response: 48679540
Conc: 46.58 ng/ml



#26 AR-1254-1


R.T.: 6.484 min
Delta R.T.: -0.001 min
Response: 45270142
Conc: 477.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.482 min
Delta R.T.: -0.005 min
Response: 31036939
Conc: 538.32 ng/ml




#27 AR-1254-2

R.T.: 6.703 min
Delta R.T.: -0.001 min
Response: 69473083
Conc: 483.93 ng/ml



#27 AR-1254-2


R.T.: 5.630 min
Delta R.T.: -0.005 min
Response: 27183704
Conc: 530.15 ng/ml



#28 AR-1254-3


R.T.: 7.071 min
Delta R.T.: -0.001 min
Response: 71457652
Conc: 487.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.033 min
Delta R.T.: -0.005 min
Response: 41354463
Conc: 509.61 ng/ml




#29 AR-1254-4

R.T.: 7.359 min
Delta R.T.: 0.000 min
Response: 52436255
Conc: 484.50 ng/ml



#29 AR-1254-4


R.T.: 6.262 min
Delta R.T.: -0.004 min
Response: 25106475
Conc: 503.40 ng/ml



#30 AR-1254-5

R.T.: 7.779 min
Delta R.T.: -0.001 min
Response: 53981924
Conc: 448.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.680 min
Delta R.T.: -0.004 min
Response: 37244480
Conc: 493.14 ng/ml