

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072522\
 Data File : P0088313.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jul 2022 18:33
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 03:42:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 15 03:57:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.447	3.607	131.8E6	52475520	51.669	55.619
2)	SA Decachlor...	10.290	8.601	93870959	48950343	44.201	46.841
Target Compounds							
26)	L6 AR-1254-1	6.482	5.483	40594553	32103261	428.153	556.816 #
27)	L6 AR-1254-2	6.702	5.631	60193144	28188232	419.292	549.742 #
28)	L6 AR-1254-3	7.069	6.033	66876848	42774669	456.570	527.108
29)	L6 AR-1254-4	7.357	6.262	50130961	25965310	463.199	520.617
30)	L6 AR-1254-5	7.777	6.680	53648619	37475162	445.759	496.193


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072522\
Data File : P0088313.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jul 2022 18:33
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 26 03:42:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 15 03:57:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.447 min
Delta R.T.: 0.000 min
Response: 131816166
Conc: 51.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.607 min
Delta R.T.: 0.002 min
Response: 52475520
Conc: 55.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.290 min
Delta R.T.: -0.009 min
Response: 93870959
Conc: 44.20 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.601 min
Delta R.T.: -0.008 min
Response: 48950343
Conc: 46.84 ng/ml



#26 AR-1254-1


R.T.: 6.482 min
Delta R.T.: -0.003 min
Response: 40594553
Conc: 428.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.483 min
Delta R.T.: -0.004 min
Response: 32103261
Conc: 556.82 ng/ml




#27 AR-1254-2

R.T.: 6.702 min
Delta R.T.: -0.002 min
Response: 60193144
Conc: 419.29 ng/ml



#27 AR-1254-2


R.T.: 5.631 min
Delta R.T.: -0.004 min
Response: 28188232
Conc: 549.74 ng/ml



#28 AR-1254-3


R.T.: 7.069 min
Delta R.T.: -0.003 min
Response: 66876848
Conc: 456.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.033 min
Delta R.T.: -0.005 min
Response: 42774669
Conc: 527.11 ng/ml




#29 AR-1254-4

R.T.: 7.357 min
Delta R.T.: -0.002 min
Response: 50130961
Conc: 463.20 ng/ml



#29 AR-1254-4


R.T.: 6.262 min
Delta R.T.: -0.004 min
Response: 25965310
Conc: 520.62 ng/ml



#30 AR-1254-5

R.T.: 7.777 min
Delta R.T.: -0.003 min
Response: 53648619
Conc: 445.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.680 min
Delta R.T.: -0.004 min
Response: 37475162
Conc: 496.19 ng/ml