

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072522\
 Data File : P0088338.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2022 04:50
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 05:39:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 15 03:57:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.445	3.605	134.9E6	51841800	52.880	54.947
2) SA Decachlor...	10.283	8.597	91483218	49620246	43.077	47.482
Target Compounds						
21) L5 AR-1248-1	5.619	4.680	26036141	11748717	487.309	512.904
22) L5 AR-1248-2	5.895	4.916	36532496	16808977	470.174	498.303
23) L5 AR-1248-3	6.101	4.957	39018093	17475835	466.113	502.854
24) L5 AR-1248-4	6.506	5.128	42847851	20174615	466.206	481.898
25) L5 AR-1248-5	6.544	5.520	42570726	19867556	473.343	551.425



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072522\
Data File : P0088338.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2022 04:50
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 26 05:39:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 15 03:57:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.445 min
Delta R.T.: -0.002 min
Response: 134906533
Conc: 52.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.605 min
Delta R.T.: 0.000 min
Response: 51841800
Conc: 54.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.283 min
Delta R.T.: -0.016 min
Response: 91483218
Conc: 43.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.597 min
Delta R.T.: -0.011 min
Response: 49620246
Conc: 47.48 ng/ml



#21 AR-1248-1


R.T.: 5.619 min
Delta R.T.: -0.005 min
Response: 26036141
Conc: 487.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.680 min
Delta R.T.: -0.006 min
Response: 11748717
Conc: 512.90 ng/ml




#22 AR-1248-2

R.T.: 5.895 min
Delta R.T.: -0.005 min
Response: 36532496
Conc: 470.17 ng/ml



#22 AR-1248-2


R.T.: 4.916 min
Delta R.T.: -0.007 min
Response: 16808977
Conc: 498.30 ng/ml



#23 AR-1248-3


R.T.: 6.101 min
Delta R.T.: -0.005 min
Response: 39018093
Conc: 466.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.957 min
Delta R.T.: -0.007 min
Response: 17475835
Conc: 502.85 ng/ml




#24 AR-1248-4

R.T.: 6.506 min
Delta R.T.: -0.005 min
Response: 42847851
Conc: 466.21 ng/ml



#24 AR-1248-4


R.T.: 5.128 min
Delta R.T.: -0.007 min
Response: 20174615
Conc: 481.90 ng/ml



#25 AR-1248-5

R.T.: 6.544 min
Delta R.T.: -0.006 min
Response: 42570726
Conc: 473.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.520 min
Delta R.T.: -0.007 min
Response: 19867556
Conc: 551.42 ng/ml