

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072621\
 Data File : P0079839.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Jul 2021 14:10
 Operator : DD\AJ
 Sample : M2940-08
 Misc : AR1221 LOD 25 PPB
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOQ-WATER-02-QT3-2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 04:49:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 23 05:30:42 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.002	4.121	928321	385638	16.416	17.346
2)	SA Decachlor...	11.116	9.535	776695	385169	19.654	22.457
Target Compounds							
8)	L2 AR-1221-1	5.255	4.386	34208	12733	57.293	49.068m
9)	L2 AR-1221-2	5.356	4.484	22551	14738	50.562	77.577m#
10)	L2 AR-1221-3	5.442	4.576	73690	33387	50.516	53.653m



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072621\
Data File : P0079839.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Jul 2021 14:10
Operator : DD\AJ
Sample : M2940-08
Misc : AR1221 LOD 25 PPB
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT3-2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 27 04:49:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072221.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 23 05:30:42 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 928321
Conc: 16.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT3-2021




#1 Tetrachloro-m-xylene

R.T.: 4.121 min
Delta R.T.: 0.000 min
Response: 385638
Conc: 17.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.116 min
Delta R.T.: 0.004 min
Response: 776695
Conc: 19.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.535 min
Delta R.T.: 0.000 min
Response: 385169
Conc: 22.46 ng/ml



#8 AR-1221-1


R.T.: 5.255 min
Delta R.T.: 0.000 min
Response: 34208
Conc: 57.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT3-2021




#8 AR-1221-1

R.T.: 4.386 min
Delta R.T.: 0.000 min
Response: 12733
Conc: 49.07 ng/ml m




#9 AR-1221-2

R.T.: 5.356 min
Delta R.T.: 0.002 min
Response: 22551
Conc: 50.56 ng/ml



#9 AR-1221-2


R.T.: 4.484 min
Delta R.T.: -0.003 min
Response: 14738
Conc: 77.58 ng/ml m



#10 AR-1221-3

R.T.: 5.442 min
Delta R.T.: 0.000 min
Response: 73690
Conc: 50.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT3-2021



#10 AR-1221-3

R.T.: 4.576 min
Delta R.T.: -0.001 min
Response: 33387
Conc: 53.65 ng/ml m