

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072623\
 Data File : P0096395.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2023 22:04
 Operator : YP/AJ
 Sample : 03756-20
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 307/308-TP-6-8-CS3

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 03:03:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 22 12:20:47 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.417	3.625	56809950	22729216	18.044	17.318
2) SA Decachlor...	10.246	8.653	36882702	19667400	18.743	18.053

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072623\
Data File : P0096395.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2023 22:04
Operator : YP/AJ
Sample : 03756-20
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
307/308-TP-6-8-CS3

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 27 03:03:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 22 12:20:47 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.417 min
Delta R.T.: 0.007 min
Response: 56809950
Conc: 18.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
307/308-TP-6-8-CS3




#1 Tetrachloro-m-xylene

R.T.: 3.625 min
Delta R.T.: -0.004 min
Response: 22729216
Conc: 17.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.246 min
Delta R.T.: 0.010 min
Response: 36882702
Conc: 18.74 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.653 min
Delta R.T.: -0.011 min
Response: 19667400
Conc: 18.05 ng/ml