

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD 0\Data\P0072718\
 Data File : P0047253.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Jul 2018 5:44
 Operator : SM/SJ
 Sample : PB111451BSD
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB111451BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 07:59:08 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 27 02:33:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.399	3.648	5662894	6476802	28.048	26.359
2) SA Decachlor...	10.090	8.641	5030936	4834458	30.846	30.755
Target Compounds						
3) L1 AR-1016-1	5.299	4.511	192162	230408	40.231	40.302
4) L1 AR-1016-2	5.649	4.926	266681	208019	45.300	39.619
5) L1 AR-1016-3	5.748	4.967	228830	177781	46.142	40.489
6) L1 AR-1016-4	6.041	5.008	205057	201852	44.085	38.927
7) L1 AR-1016-5	6.182	5.180	156857	260568	30.090	44.423 #
8) L2 AR-1221-1	4.602	3.858	18692	24034	8.452	10.185
9) L2 AR-1221-2	4.703	3.947	100683	111102	61.573	61.245
10) L2 AR-1221-3	4.773	4.023	216147	138588	41.866	23.975 #
11) L3 AR-1232-1	5.105	4.320	185299	360124	86.165	153.131 #
12) L3 AR-1232-2	5.566	4.731	274539	286480	93.302	93.747
13) L3 AR-1232-3	5.588	4.750	388443	440292	90.415	95.347
14) L3 AR-1232-4	5.649	4.805	266681	239521	96.348	77.480
15) L3 AR-1232-5	5.748	4.967	228830	177781	102.471	99.335
16) L4 AR-1242-1	5.588	4.750	388443	440292	52.849	54.980
17) L4 AR-1242-2	5.649	4.926	266681	208019	57.619	50.494
18) L4 AR-1242-3	5.748	5.008	228830	201852	59.560	48.131
19) L4 AR-1242-4	6.041	5.180	205057	260568	53.341	55.179
20) L4 AR-1242-5	6.182	5.327	156857	234829	36.644	60.653 #
21) L5 AR-1248-1	6.041	5.180	205057	260568	32.793	34.928
22) L5 AR-1248-2	6.182	5.327	156857	234829	28.794	43.893 #
23) L5 AR-1248-3	6.444	5.531	31033	169243	4.410	17.009 #
24) L5 AR-1248-4	6.483	5.560	37732	76461	5.621	10.910 #
25) L5 AR-1248-5	6.634	6.094	162451	370806	22.953	77.806 #
26) L6 AR-1254-1	6.634	5.531	162451	169243	13.284	13.754
27) L6 AR-1254-2	6.915	5.676	49641	178967	6.656	17.192 #
28) L6 AR-1254-3	7.002	6.094	117875	370806	9.039	21.365 #
29) L6 AR-1254-4	7.285	6.306	60542	33839	6.212	3.123 #
30) L6 AR-1254-5	7.547	6.630	175381	115321	23.739	15.079 #
31) L7 AR-1260-1	7.163	6.210	375408	448017	40.035	40.544
32) L7 AR-1260-2	7.420	6.395	429146	543562	38.484	40.540
33) L7 AR-1260-3	7.702	6.723	542632	626912	42.175	43.123
34) L7 AR-1260-4	8.317	7.260	649199	903859	36.497	41.078
35) L7 AR-1260-5	8.639	7.606	462951	1361041	40.541	87.248 #
36) L8 AR-1262-1	7.163	6.395	375408	543562	45.314	47.941
37) L8 AR-1262-2	7.420	6.550	429146	521859	44.280	46.246
38) L8 AR-1262-3	7.777	6.761	314091	407606	25.593	27.008

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD 0\Data\P0072718\
 Data File : P0047253.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Jul 2018 5:44
 Operator : SM/SJ
 Sample : PB111451BSD
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB111451BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 07:59:08 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 27 02:33:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
39)	L8 AR-1262-4	8.002	7.021	365365	341414	31.027	25.927
40)	L8 AR-1262-5	8.317	7.260	649199	903859	30.216	34.994
41)	L9 AR-1268-1	8.639	7.546	462951	315534	19.947	12.136 #
42)	L9 AR-1268-2	8.693	7.606	290150	1361041	13.542	54.632 #
43)	L9 AR-1268-3	8.934	7.815	42448	61138	2.351	3.098 #
44)	L9 AR-1268-4	9.357	8.099	235815	236653	29.573	28.211
45)	L9 AR-1268-5	9.760	8.387	102877	171540	1.888	3.142 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0072718\
Data File : P0047253.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Jul 2018 5:44
Operator : SM/SJ
Sample : PB111451BSD
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB111451BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 27 07:59:08 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072718.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 27 02:33:29 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.399 min
Delta R.T.: 0.001 min
Response: 5662894
Conc: 28.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#1 Tetrachloro-m-xylene

R.T.: 3.648 min
Delta R.T.: 0.002 min
Response: 6476802
Conc: 26.36 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.090 min
Delta R.T.: 0.000 min
Response: 5030936
Conc: 30.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.641 min
Delta R.T.: -0.003 min
Response: 4834458
Conc: 30.76 ng/ml



#3 AR-1016-1


R.T.: 5.299 min
Delta R.T.: 0.000 min
Response: 192162
Conc: 40.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#3 AR-1016-1

R.T.: 4.511 min
Delta R.T.: 0.000 min
Response: 230408
Conc: 40.30 ng/ml




#4 AR-1016-2

R.T.: 5.649 min
Delta R.T.: -0.001 min
Response: 266681
Conc: 45.30 ng/ml



#4 AR-1016-2


R.T.: 4.926 min
Delta R.T.: 0.000 min
Response: 208019
Conc: 39.62 ng/ml



#5 AR-1016-3


R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 228830
Conc: 46.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#5 AR-1016-3

R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 177781
Conc: 40.49 ng/ml




#6 AR-1016-4

R.T.: 6.041 min
Delta R.T.: 0.000 min
Response: 205057
Conc: 44.09 ng/ml



#6 AR-1016-4


R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 201852
Conc: 38.93 ng/ml



#7 AR-1016-5


R.T.: 6.182 min
Delta R.T.: -0.002 min
Response: 156857
Conc: 30.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#7 AR-1016-5

R.T.: 5.180 min
Delta R.T.: 0.000 min
Response: 260568
Conc: 44.42 ng/ml




#8 AR-1221-1

R.T.: 4.602 min
Delta R.T.: -0.006 min
Response: 18692
Conc: 8.45 ng/ml



#8 AR-1221-1


R.T.: 3.858 min
Delta R.T.: -0.002 min
Response: 24034
Conc: 10.19 ng/ml



#9 AR-1221-2


R.T.: 4.703 min
Delta R.T.: 0.009 min
Response: 100683
Conc: 61.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#9 AR-1221-2

R.T.: 3.947 min
Delta R.T.: 0.000 min
Response: 111102
Conc: 61.24 ng/ml




#10 AR-1221-3

R.T.: 4.773 min
Delta R.T.: 0.001 min
Response: 216147
Conc: 41.87 ng/ml



#10 AR-1221-3


R.T.: 4.023 min
Delta R.T.: 0.000 min
Response: 138588
Conc: 23.97 ng/ml



#11 AR-1232-1


R.T.: 5.105 min
Delta R.T.: 0.000 min
Response: 185299
Conc: 86.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#11 AR-1232-1

R.T.: 4.320 min
Delta R.T.: 0.000 min
Response: 360124
Conc: 153.13 ng/ml




#12 AR-1232-2

R.T.: 5.566 min
Delta R.T.: 0.000 min
Response: 274539
Conc: 93.30 ng/ml



#12 AR-1232-2


R.T.: 4.731 min
Delta R.T.: 0.001 min
Response: 286480
Conc: 93.75 ng/ml



#13 AR-1232-3


R.T.: 5.588 min
Delta R.T.: 0.000 min
Response: 388443
Conc: 90.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#13 AR-1232-3

R.T.: 4.750 min
Delta R.T.: 0.000 min
Response: 440292
Conc: 95.35 ng/ml




#14 AR-1232-4

R.T.: 5.649 min
Delta R.T.: 0.000 min
Response: 266681
Conc: 96.35 ng/ml



#14 AR-1232-4


R.T.: 4.805 min
Delta R.T.: 0.000 min
Response: 239521
Conc: 77.48 ng/ml



#15 AR-1232-5


R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 228830
Conc: 102.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#15 AR-1232-5

R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 177781
Conc: 99.33 ng/ml




#16 AR-1242-1

R.T.: 5.588 min
Delta R.T.: 0.000 min
Response: 388443
Conc: 52.85 ng/ml



#16 AR-1242-1


R.T.: 4.750 min
Delta R.T.: 0.000 min
Response: 440292
Conc: 54.98 ng/ml



#17 AR-1242-2


R.T.: 5.649 min
Delta R.T.: -0.001 min
Response: 266681
Conc: 57.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#17 AR-1242-2

R.T.: 4.926 min
Delta R.T.: 0.000 min
Response: 208019
Conc: 50.49 ng/ml




#18 AR-1242-3

R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 228830
Conc: 59.56 ng/ml



#18 AR-1242-3


R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 201852
Conc: 48.13 ng/ml



#19 AR-1242-4


R.T.: 6.041 min
Delta R.T.: 0.000 min
Response: 205057
Conc: 53.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#19 AR-1242-4

R.T.: 5.180 min
Delta R.T.: 0.000 min
Response: 260568
Conc: 55.18 ng/ml




#20 AR-1242-5

R.T.: 6.182 min
Delta R.T.: -0.001 min
Response: 156857
Conc: 36.64 ng/ml



#20 AR-1242-5


R.T.: 5.327 min
Delta R.T.: -0.002 min
Response: 234829
Conc: 60.65 ng/ml



#21 AR-1248-1


R.T.: 6.041 min
Delta R.T.: -0.002 min
Response: 205057
Conc: 32.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#21 AR-1248-1

R.T.: 5.180 min
Delta R.T.: 0.000 min
Response: 260568
Conc: 34.93 ng/ml




#22 AR-1248-2

R.T.: 6.182 min
Delta R.T.: 0.000 min
Response: 156857
Conc: 28.79 ng/ml



#22 AR-1248-2


R.T.: 5.327 min
Delta R.T.: -0.001 min
Response: 234829
Conc: 43.89 ng/ml



#23 AR-1248-3


R.T.: 6.444 min
Delta R.T.: -0.001 min
Response: 31033
Conc: 4.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#23 AR-1248-3

R.T.: 5.531 min
Delta R.T.: 0.000 min
Response: 169243
Conc: 17.01 ng/ml




#24 AR-1248-4

R.T.: 6.483 min
Delta R.T.: 0.000 min
Response: 37732
Conc: 5.62 ng/ml



#24 AR-1248-4


R.T.: 5.560 min
Delta R.T.: -0.012 min
Response: 76461
Conc: 10.91 ng/ml



#25 AR-1248-5


R.T.: 6.634 min
Delta R.T.: -0.005 min
Response: 162451
Conc: 22.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#25 AR-1248-5

R.T.: 6.094 min
Delta R.T.: 0.013 min
Response: 370806
Conc: 77.81 ng/ml




#26 AR-1254-1

R.T.: 6.634 min
Delta R.T.: -0.001 min
Response: 162451
Conc: 13.28 ng/ml



#26 AR-1254-1


R.T.: 5.531 min
Delta R.T.: 0.000 min
Response: 169243
Conc: 13.75 ng/ml



#27 AR-1254-2


R.T.: 6.915 min
Delta R.T.: 0.000 min
Response: 49641
Conc: 6.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#27 AR-1254-2

R.T.: 5.676 min
Delta R.T.: -0.001 min
Response: 178967
Conc: 17.19 ng/ml




#28 AR-1254-3

R.T.: 7.002 min
Delta R.T.: 0.000 min
Response: 117875
Conc: 9.04 ng/ml



#28 AR-1254-3


R.T.: 6.094 min
Delta R.T.: 0.012 min
Response: 370806
Conc: 21.37 ng/ml



#29 AR-1254-4


R.T.: 7.285 min
Delta R.T.: -0.002 min
Response: 60542
Conc: 6.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#29 AR-1254-4

R.T.: 6.306 min
Delta R.T.: -0.002 min
Response: 33839
Conc: 3.12 ng/ml




#30 AR-1254-5

R.T.: 7.547 min
Delta R.T.: -0.011 min
Response: 175381
Conc: 23.74 ng/ml



#30 AR-1254-5


R.T.: 6.630 min
Delta R.T.: 0.007 min
Response: 115321
Conc: 15.08 ng/ml



#31 AR-1260-1


R.T.: 7.163 min
Delta R.T.: -0.002 min
Response: 375408
Conc: 40.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#31 AR-1260-1

R.T.: 6.210 min
Delta R.T.: -0.001 min
Response: 448017
Conc: 40.54 ng/ml




#32 AR-1260-2

R.T.: 7.420 min
Delta R.T.: -0.001 min
Response: 429146
Conc: 38.48 ng/ml



#32 AR-1260-2


R.T.: 6.395 min
Delta R.T.: -0.002 min
Response: 543562
Conc: 40.54 ng/ml



#33 AR-1260-3


R.T.: 7.702 min
Delta R.T.: -0.001 min
Response: 542632
Conc: 42.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#33 AR-1260-3

R.T.: 6.723 min
Delta R.T.: -0.001 min
Response: 626912
Conc: 43.12 ng/ml




#34 AR-1260-4

R.T.: 8.317 min
Delta R.T.: -0.001 min
Response: 649199
Conc: 36.50 ng/ml



#34 AR-1260-4


R.T.: 7.260 min
Delta R.T.: -0.002 min
Response: 903859
Conc: 41.08 ng/ml



#35 AR-1260-5


R.T.: 8.639 min
Delta R.T.: -0.001 min
Response: 462951
Conc: 40.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#35 AR-1260-5

R.T.: 7.606 min
Delta R.T.: -0.003 min
Response: 1361041
Conc: 87.25 ng/ml




#36 AR-1262-1

R.T.: 7.163 min
Delta R.T.: -0.001 min
Response: 375408
Conc: 45.31 ng/ml



#36 AR-1262-1


R.T.: 6.395 min
Delta R.T.: -0.001 min
Response: 543562
Conc: 47.94 ng/ml



#37 AR-1262-2


R.T.: 7.420 min
Delta R.T.: 0.000 min
Response: 429146
Conc: 44.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#37 AR-1262-2

R.T.: 6.550 min
Delta R.T.: -0.003 min
Response: 521859
Conc: 46.25 ng/ml




#38 AR-1262-3

R.T.: 7.777 min
Delta R.T.: 0.000 min
Response: 314091
Conc: 25.59 ng/ml



#38 AR-1262-3


R.T.: 6.761 min
Delta R.T.: -0.001 min
Response: 407606
Conc: 27.01 ng/ml



#39 AR-1262-4


R.T.: 8.002 min
Delta R.T.: -0.001 min
Response: 365365
Conc: 31.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#39 AR-1262-4

R.T.: 7.021 min
Delta R.T.: -0.002 min
Response: 341414
Conc: 25.93 ng/ml




#40 AR-1262-5

R.T.: 8.317 min
Delta R.T.: 0.000 min
Response: 649199
Conc: 30.22 ng/ml



#40 AR-1262-5


R.T.: 7.260 min
Delta R.T.: -0.001 min
Response: 903859
Conc: 34.99 ng/ml



#41 AR-1268-1


R.T.: 8.639 min
Delta R.T.: 0.016 min
Response: 462951
Conc: 19.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#41 AR-1268-1

R.T.: 7.546 min
Delta R.T.: 0.000 min
Response: 315534
Conc: 12.14 ng/ml




#42 AR-1268-2

R.T.: 8.693 min
Delta R.T.: -0.023 min
Response: 290150
Conc: 13.54 ng/ml



#42 AR-1268-2


R.T.: 7.606 min
Delta R.T.: -0.003 min
Response: 1361041
Conc: 54.63 ng/ml



#43 AR-1268-3


R.T.: 8.934 min
Delta R.T.: -0.003 min
Response: 42448
Conc: 2.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD




#43 AR-1268-3

R.T.: 7.815 min
Delta R.T.: -0.002 min
Response: 61138
Conc: 3.10 ng/ml




#44 AR-1268-4

R.T.: 9.357 min
Delta R.T.: 0.000 min
Response: 235815
Conc: 29.57 ng/ml



#44 AR-1268-4


R.T.: 8.099 min
Delta R.T.: -0.002 min
Response: 236653
Conc: 28.21 ng/ml



#45 AR-1268-5

R.T.: 9.760 min
Delta R.T.: -0.003 min
Response: 102877
Conc: 1.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB111451BSD



#45 AR-1268-5

R.T.: 8.387 min
Delta R.T.: -0.003 min
Response: 171540
Conc: 3.14 ng/ml