

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072723\
 Data File : P0096433.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jul 2023 19:30
 Operator : YP/AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 28 01:11:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 28 01:10:11 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.418	3.627	85897216	36435182	25.210	24.995
2)	SA Decachlor...	10.253	8.655	99742889	54828082	25.715	26.338
Target Compounds							
41)	L9 AR-1268-1	8.710	7.548	78574714	48800445	255.264	261.214
42)	L9 AR-1268-2	8.805	7.614	69527632	44086947	250.769	259.906
43)	L9 AR-1268-3	9.039	7.820	64990282	38410706	253.572	262.243
44)	L9 AR-1268-4	9.475	8.110	24167150	15762181	251.459	262.340
45)	L9 AR-1268-5	9.903	8.401	208.1E6	111.6E6	251.725	253.017



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072723\
Data File : P0096433.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2023 19:30
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 01:11:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072723.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 28 01:10:11 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.418 min
Delta R.T.: 0.000 min
Response: 85897216
Conc: 25.21 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.627 min
Delta R.T.: 0.000 min
Response: 36435182
Conc: 24.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.253 min
Delta R.T.: 0.000 min
Response: 99742889
Conc: 25.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.655 min
Delta R.T.: 0.000 min
Response: 54828082
Conc: 26.34 ng/ml



#41 AR-1268-1


R.T.: 8.710 min
Delta R.T.: 0.000 min
Response: 78574714
Conc: 255.26 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.548 min
Delta R.T.: 0.000 min
Response: 48800445
Conc: 261.21 ng/ml




#42 AR-1268-2

R.T.: 8.805 min
Delta R.T.: 0.000 min
Response: 69527632
Conc: 250.77 ng/ml



#42 AR-1268-2


R.T.: 7.614 min
Delta R.T.: 0.000 min
Response: 44086947
Conc: 259.91 ng/ml



#43 AR-1268-3


R.T.: 9.039 min
Delta R.T.: 0.000 min
Response: 64990282
Conc: 253.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.820 min
Delta R.T.: 0.000 min
Response: 38410706
Conc: 262.24 ng/ml




#44 AR-1268-4

R.T.: 9.475 min
Delta R.T.: 0.000 min
Response: 24167150
Conc: 251.46 ng/ml



#44 AR-1268-4


R.T.: 8.110 min
Delta R.T.: 0.000 min
Response: 15762181
Conc: 262.34 ng/ml



#45 AR-1268-5

R.T.: 9.903 min
Delta R.T.: 0.000 min
Response: 208120367
Conc: 251.72 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.401 min
Delta R.T.: 0.000 min
Response: 111615922
Conc: 253.02 ng/ml