

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072723\
 Data File : P0096418.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jul 2023 15:15
 Operator : YP/AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/27/2023
 Supervised By :Ankita Jodhani 11/27/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 18:03:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 27 18:01:34 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.418	3.643	15944268	6639091	4.621	4.522
2) SA Decachlor...	10.252	8.671	10347864	6297112	4.522	5.273
Target Compounds						
16) L4 AR-1242-1	5.601	4.731	3982749	2000528	48.195	50.496
17) L4 AR-1242-2	5.623	4.750	5516683	2646022	46.741	48.329m
18) L4 AR-1242-3	5.686	4.927	3395154	1398860	46.289	45.990
19) L4 AR-1242-4	5.785	5.011	2790504	1531886	47.120	47.340
20) L4 AR-1242-5	6.530	5.535	2700253	1720233	49.349	46.224

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072723\
Data File : P0096418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2023 15:15
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/27/2023
Supervised By :Ankita Jodhani 11/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 27 18:03:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072723.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 27 18:01:34 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

