

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0072821\  
 Data File : P0079877.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 28 Jul 2021 18:20  
 Operator : DD\AJ  
 Sample : AR1242ICV500  
 Misc :  
 ALS Vial : 28 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 ICVPO072821AR1242

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jul 28 18:58:38 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0072821.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Wed Jul 28 18:45:36 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound        | RT#1   | RT#2   | Resp#1  | Resp#2  | ng/ml    | ng/ml     |
|-----------------------------|-----------------|--------|--------|---------|---------|----------|-----------|
| -----                       |                 |        |        |         |         |          |           |
| System Monitoring Compounds |                 |        |        |         |         |          |           |
| 1)                          | SA Tetrachlo... | 5.001  | 4.120  | 3194497 | 1196228 | 53.014   | 52.234    |
| 2)                          | SA Decachlor... | 11.113 | 9.530  | 2520183 | 1070218 | 53.724   | 51.498    |
| Target Compounds            |                 |        |        |         |         |          |           |
| 3)                          | L1 AR-1016-1    | 6.332  | 5.399  | 861488  | 362209  | 391.899  | 385.361   |
| 4)                          | L1 AR-1016-2    | 6.356  | 5.419  | 1184328 | 485298  | 388.173  | 381.459   |
| 5)                          | L1 AR-1016-3    | 6.422  | 5.615  | 731240  | 256288  | 388.582  | 363.519   |
| 6)                          | L1 AR-1016-4    | 6.531  | 5.666  | 601590  | 209421  | 390.165  | 346.572   |
| 7)                          | L1 AR-1016-5    | 6.847  | 5.900  | 782701  | 286298  | 517.346  | 388.754   |
| 8)                          | L2 AR-1221-1    | 5.248  | 4.382  | 90478   | 39104   | 140.412  | 135.417   |
| 9)                          | L2 AR-1221-2    | 5.354  | 4.484  | 123038  | 56900   | 267.955  | 265.302   |
| 10)                         | L2 AR-1221-3    | 5.442  | 4.575  | 472189  | 201532  | 306.008  | 310.424   |
| 11)                         | L3 AR-1232-1    | 5.442  | 4.575  | 472189  | 201532  | 331.347  | 334.213   |
| 12)                         | L3 AR-1232-2    | 6.035  | 5.419  | 612725  | 485298  | 877.144  | 862.991   |
| 13)                         | L3 AR-1232-3    | 6.356  | 5.615  | 1184328 | 256288  | 888.616  | 851.279   |
| 14)                         | L3 AR-1232-4    | 6.531  | 5.711  | 601590  | 268865  | 923.743  | 944.480   |
| 15)                         | L3 AR-1232-5    | 6.629  | 5.900  | 460840  | 286298  | 1031.989 | 953.334   |
| 16)                         | L4 AR-1242-1    | 6.332  | 5.399  | 861488  | 362209  | 516.886  | 508.590   |
| 17)                         | L4 AR-1242-2    | 6.356  | 5.419  | 1184328 | 485298  | 514.966  | 501.152   |
| 18)                         | L4 AR-1242-3    | 6.422  | 5.615  | 731240  | 256288  | 515.107  | 477.650   |
| 19)                         | L4 AR-1242-4    | 6.531  | 5.711  | 601590  | 268865  | 530.149  | 478.817   |
| 20)                         | L4 AR-1242-5    | 7.317  | 6.284  | 639400  | 343708  | 503.630  | 528.219   |
| 21)                         | L5 AR-1248-1    | 6.332  | 5.399  | 861488  | 362209  | 679.576  | 655.940   |
| 22)                         | L5 AR-1248-2    | 6.629  | 5.666  | 460840  | 209421  | 264.342  | 254.187   |
| 23)                         | L5 AR-1248-3    | 6.847  | 5.711  | 782701  | 268865  | 381.896  | 323.975   |
| 24)                         | L5 AR-1248-4    | 7.278  | 5.900  | 611921  | 286298  | 261.405  | 302.985   |
| 25)                         | L5 AR-1248-5    | 7.317  | 6.328  | 639400  | 324258  | 281.177  | 353.650 # |
| 26)                         | L6 AR-1254-1    | 7.247  | 6.284  | 476894  | 343708  | 218.319  | 253.459   |
| 27)                         | L6 AR-1254-2    | 7.483  | 6.444  | 541059  | 174348  | 164.768  | 145.405   |
| 28)                         | L6 AR-1254-3    | 7.861  | 6.873  | 214916  | 107913  | 62.418   | 57.675    |
| 29)                         | L6 AR-1254-4    | 8.158  | 7.114  | 155774  | 79810   | 61.247   | 67.413    |
| 30)                         | L6 AR-1254-5    | 8.589  | 7.550  | 36942   | 27910   | 13.891   | 17.423 #  |
| 31)                         | L7 AR-1260-1    | 0.000  | 7.024  | 0       | 40297   | N.D.     | 31.133 #  |
| 32)                         | L7 AR-1260-2    | 8.280  | 7.208  | 1153    | 18022   | 0.375    | 11.781 #  |
| 33)                         | L7 AR-1260-3    | 0.000  | 7.366  | 0       | 31009   | N.D.     | 21.238 #  |
| 36)                         | L8 AR-1262-1    | 0.000  | 7.550  | 0       | 27910   | N.D.     | 33.157 #  |
| 38)                         | L8 AR-1262-3    | 9.554  | 8.375f | 22814   | 26662   | 6.032    | 24.005 #  |
| 39)                         | L8 AR-1262-4    | 0.000  | 8.456  | 0       | 32365   | N.D.     | 16.187 #  |
| 41)                         | L9 AR-1268-1    | 9.554  | 8.375f | 22814   | 26662   | 3.524    | 8.775 #   |
| 42)                         | L9 AR-1268-2    | 0.000  | 8.456  | 0       | 32365   | N.D.     | 11.895 #  |
| 43)                         | L9 AR-1268-3    | 0.000  | 8.654f | 0       | 24482   | N.D.     | 10.561 #  |
| -----                       |                 |        |        |         |         |          |           |

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0072821\  
Data File : P0079877.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 28 Jul 2021 18:20  
Operator : DD\AJ  
Sample : AR1242ICV500  
Misc :  
ALS Vial : 28 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jul 28 18:58:38 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0072821.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Jul 28 18:45:36 2021  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm


| Compound                                                              | RT#1 | RT#2 | Resp#1 | Resp#2 | ng/ml | ng/ml |
|-----------------------------------------------------------------------|------|------|--------|--------|-------|-------|
| -----                                                                 |      |      |        |        |       |       |
| (f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int. |      |      |        |        |       |       |


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\PO072821\  
Data File : PO079877.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 28 Jul 2021 18:20  
Operator : DD\AJ  
Sample : AR1242ICV500  
Misc :  
ALS Vial : 28 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jul 28 18:58:38 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\PO072821.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Wed Jul 28 18:45:36 2021  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.001 min  
Delta R.T.: 0.001 min  
Response: 3194497  
Conc: 53.01 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#1 Tetrachloro-m-xylene

R.T.: 4.120 min  
Delta R.T.: 0.000 min  
Response: 1196228  
Conc: 52.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.113 min  
Delta R.T.: 0.003 min  
Response: 2520183  
Conc: 53.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.530 min  
Delta R.T.: 0.000 min  
Response: 1070218  
Conc: 51.50 ng/ml



#3 AR-1016-1


R.T.: 6.332 min  
Delta R.T.: 0.000 min  
Response: 861488  
Conc: 391.90 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#3 AR-1016-1

R.T.: 5.399 min  
Delta R.T.: 0.000 min  
Response: 362209  
Conc: 385.36 ng/ml




#4 AR-1016-2

R.T.: 6.356 min  
Delta R.T.: 0.000 min  
Response: 1184328  
Conc: 388.17 ng/ml



#4 AR-1016-2


R.T.: 5.419 min  
Delta R.T.: 0.000 min  
Response: 485298  
Conc: 381.46 ng/ml



#5 AR-1016-3


R.T.: 6.422 min  
Delta R.T.: 0.000 min  
Response: 731240  
Conc: 388.58 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#5 AR-1016-3

R.T.: 5.615 min  
Delta R.T.: 0.000 min  
Response: 256288  
Conc: 363.52 ng/ml




#6 AR-1016-4

R.T.: 6.531 min  
Delta R.T.: 0.002 min  
Response: 601590  
Conc: 390.17 ng/ml



#6 AR-1016-4


R.T.: 5.666 min  
Delta R.T.: 0.002 min  
Response: 209421  
Conc: 346.57 ng/ml



#7 AR-1016-5


R.T.: 6.847 min  
Delta R.T.: 0.002 min  
Response: 782701  
Conc: 517.35 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#7 AR-1016-5

R.T.: 5.900 min  
Delta R.T.: 0.000 min  
Response: 286298  
Conc: 388.75 ng/ml




#8 AR-1221-1

R.T.: 5.248 min  
Delta R.T.: -0.006 min  
Response: 90478  
Conc: 140.41 ng/ml



#8 AR-1221-1


R.T.: 4.382 min  
Delta R.T.: -0.002 min  
Response: 39104  
Conc: 135.42 ng/ml



#9 AR-1221-2


R.T.: 5.354 min  
Delta R.T.: 0.000 min  
Response: 123038  
Conc: 267.96 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#9 AR-1221-2

R.T.: 4.484 min  
Delta R.T.: -0.002 min  
Response: 56900  
Conc: 265.30 ng/ml




#10 AR-1221-3

R.T.: 5.442 min  
Delta R.T.: 0.000 min  
Response: 472189  
Conc: 306.01 ng/ml



#10 AR-1221-3


R.T.: 4.575 min  
Delta R.T.: 0.000 min  
Response: 201532  
Conc: 310.42 ng/ml



#11 AR-1232-1


R.T.: 5.442 min  
Delta R.T.: 0.001 min  
Response: 472189  
Conc: 331.35 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#11 AR-1232-1

R.T.: 4.575 min  
Delta R.T.: 0.000 min  
Response: 201532  
Conc: 334.21 ng/ml




#12 AR-1232-2

R.T.: 6.035 min  
Delta R.T.: 0.000 min  
Response: 612725  
Conc: 877.14 ng/ml



#12 AR-1232-2


R.T.: 5.419 min  
Delta R.T.: 0.000 min  
Response: 485298  
Conc: 862.99 ng/ml



#13 AR-1232-3


R.T.: 6.356 min  
Delta R.T.: 0.000 min  
Response: 1184328  
Conc: 888.62 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#13 AR-1232-3

R.T.: 5.615 min  
Delta R.T.: 0.000 min  
Response: 256288  
Conc: 851.28 ng/ml




#14 AR-1232-4

R.T.: 6.531 min  
Delta R.T.: 0.000 min  
Response: 601590  
Conc: 923.74 ng/ml



#14 AR-1232-4


R.T.: 5.711 min  
Delta R.T.: 0.000 min  
Response: 268865  
Conc: 944.48 ng/ml



#15 AR-1232-5


R.T.: 6.629 min  
Delta R.T.: 0.001 min  
Response: 460840  
Conc: 1031.99 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#15 AR-1232-5

R.T.: 5.900 min  
Delta R.T.: 0.000 min  
Response: 286298  
Conc: 953.33 ng/ml




#16 AR-1242-1

R.T.: 6.332 min  
Delta R.T.: 0.000 min  
Response: 861488  
Conc: 516.89 ng/ml



#16 AR-1242-1


R.T.: 5.399 min  
Delta R.T.: 0.000 min  
Response: 362209  
Conc: 508.59 ng/ml



#17 AR-1242-2


R.T.: 6.356 min  
Delta R.T.: 0.000 min  
Response: 1184328  
Conc: 514.97 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#17 AR-1242-2

R.T.: 5.419 min  
Delta R.T.: 0.000 min  
Response: 485298  
Conc: 501.15 ng/ml




#18 AR-1242-3

R.T.: 6.422 min  
Delta R.T.: 0.001 min  
Response: 731240  
Conc: 515.11 ng/ml



#18 AR-1242-3


R.T.: 5.615 min  
Delta R.T.: 0.000 min  
Response: 256288  
Conc: 477.65 ng/ml



#19 AR-1242-4


R.T.: 6.531 min  
Delta R.T.: 0.001 min  
Response: 601590  
Conc: 530.15 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#19 AR-1242-4

R.T.: 5.711 min  
Delta R.T.: 0.000 min  
Response: 268865  
Conc: 478.82 ng/ml




#20 AR-1242-5

R.T.: 7.317 min  
Delta R.T.: 0.000 min  
Response: 639400  
Conc: 503.63 ng/ml



#20 AR-1242-5


R.T.: 6.284 min  
Delta R.T.: 0.000 min  
Response: 343708  
Conc: 528.22 ng/ml



#21 AR-1248-1


R.T.: 6.332 min  
Delta R.T.: 0.000 min  
Response: 861488  
Conc: 679.58 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#21 AR-1248-1

R.T.: 5.399 min  
Delta R.T.: 0.000 min  
Response: 362209  
Conc: 655.94 ng/ml




#22 AR-1248-2

R.T.: 6.629 min  
Delta R.T.: 0.001 min  
Response: 460840  
Conc: 264.34 ng/ml



#22 AR-1248-2


R.T.: 5.666 min  
Delta R.T.: 0.001 min  
Response: 209421  
Conc: 254.19 ng/ml



#23 AR-1248-3


R.T.: 6.847 min  
Delta R.T.: 0.001 min  
Response: 782701  
Conc: 381.90 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#23 AR-1248-3

R.T.: 5.711 min  
Delta R.T.: 0.000 min  
Response: 268865  
Conc: 323.98 ng/ml




#24 AR-1248-4

R.T.: 7.278 min  
Delta R.T.: 0.000 min  
Response: 611921  
Conc: 261.40 ng/ml



#24 AR-1248-4


R.T.: 5.900 min  
Delta R.T.: 0.001 min  
Response: 286298  
Conc: 302.98 ng/ml



#25 AR-1248-5


R.T.: 7.317 min  
Delta R.T.: 0.000 min  
Response: 639400  
Conc: 281.18 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#25 AR-1248-5

R.T.: 6.328 min  
Delta R.T.: 0.000 min  
Response: 324258  
Conc: 353.65 ng/ml




#26 AR-1254-1

R.T.: 7.247 min  
Delta R.T.: 0.000 min  
Response: 476894  
Conc: 218.32 ng/ml



#26 AR-1254-1


R.T.: 6.284 min  
Delta R.T.: 0.000 min  
Response: 343708  
Conc: 253.46 ng/ml



#27 AR-1254-2


R.T.: 7.483 min  
Delta R.T.: 0.006 min  
Response: 541059  
Conc: 164.77 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#27 AR-1254-2

R.T.: 6.444 min  
Delta R.T.: 0.000 min  
Response: 174348  
Conc: 145.40 ng/ml




#28 AR-1254-3

R.T.: 7.861 min  
Delta R.T.: 0.000 min  
Response: 214916  
Conc: 62.42 ng/ml



#28 AR-1254-3


R.T.: 6.873 min  
Delta R.T.: 0.000 min  
Response: 107913  
Conc: 57.67 ng/ml



#29 AR-1254-4


R.T.: 8.158 min  
Delta R.T.: 0.000 min  
Response: 155774  
Conc: 61.25 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#29 AR-1254-4

R.T.: 7.114 min  
Delta R.T.: 0.000 min  
Response: 79810  
Conc: 67.41 ng/ml




#30 AR-1254-5

R.T.: 8.589 min  
Delta R.T.: 0.002 min  
Response: 36942  
Conc: 13.89 ng/ml



#30 AR-1254-5


R.T.: 7.550 min  
Delta R.T.: 0.001 min  
Response: 27910  
Conc: 17.42 ng/ml



#31 AR-1260-1


R.T.: 0.000 min  
Exp R.T.: 8.027 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#31 AR-1260-1

R.T.: 7.024 min  
Delta R.T.: 0.014 min  
Response: 40297  
Conc: 31.13 ng/ml




#32 AR-1260-2

R.T.: 8.280 min  
Delta R.T.: -0.014 min  
Response: 1153  
Conc: 0.38 ng/ml



#32 AR-1260-2


R.T.: 7.208 min  
Delta R.T.: 0.000 min  
Response: 18022  
Conc: 11.78 ng/ml



#33 AR-1260-3


R.T.: 0.000 min  
Exp R.T.: 8.661 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#33 AR-1260-3

R.T.: 7.366 min  
Delta R.T.: 0.000 min  
Response: 31009  
Conc: 21.24 ng/ml




#36 AR-1262-1

R.T.: 0.000 min  
Exp R.T.: 8.661 min  
Response: 0  
Conc: N.D.



#36 AR-1262-1


R.T.: 7.550 min  
Delta R.T.: 0.001 min  
Response: 27910  
Conc: 33.16 ng/ml



#38 AR-1262-3


R.T.: 9.554 min  
Delta R.T.: 0.007 min  
Response: 22814  
Conc: 6.03 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#38 AR-1262-3

R.T.: 8.375 min  
Delta R.T.: -0.022 min  
Response: 26662  
Conc: 24.00 ng/ml




#39 AR-1262-4

R.T.: 0.000 min  
Exp R.T. : 9.645 min  
Response: 0  
Conc: N.D.



#39 AR-1262-4


R.T.: 8.456 min  
Delta R.T.: -0.006 min  
Response: 32365  
Conc: 16.19 ng/ml



#41 AR-1268-1


R.T.: 9.554 min  
Delta R.T.: 0.008 min  
Response: 22814  
Conc: 3.52 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242




#41 AR-1268-1

R.T.: 8.375 min  
Delta R.T.: -0.022 min  
Response: 26662  
Conc: 8.77 ng/ml




#42 AR-1268-2

R.T.: 0.000 min  
Exp R.T. : 9.646 min  
Response: 0  
Conc: N.D.



#42 AR-1268-2


R.T.: 8.456 min  
Delta R.T.: -0.007 min  
Response: 32365  
Conc: 11.89 ng/ml



#43 AR-1268-3

R.T.: 0.000 min  
Exp R.T.: 9.878 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO072821AR1242



#43 AR-1268-3

R.T.: 8.654 min  
Delta R.T.: -0.020 min  
Response: 24482  
Conc: 10.56 ng/ml