

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072921\
 Data File : P0079975.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Jul 2021 1:03
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 03:07:48 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 28 18:45:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.003	4.120	3447901	1313972	57.219	57.376
2)	SA Decachlor...	11.117	9.530	2452638	1080059	52.284	51.972
Target Compounds							
26)	L6 AR-1254-1	7.249	6.284	1200825	740302	549.730	545.917
27)	L6 AR-1254-2	7.479	6.445	1759210	648643	535.731	540.962
28)	L6 AR-1254-3	7.862	6.872	1754924	1019639	509.685	544.950
29)	L6 AR-1254-4	8.160	7.114	1341743	631727	527.543	533.603
30)	L6 AR-1254-5	8.590	7.549	1335568	849759	502.197	530.471



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072921\
Data File : P0079975.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Jul 2021 1:03
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 03:07:48 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072821.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 28 18:45:36 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.003 min
Delta R.T.: 0.003 min
Response: 3447901
Conc: 57.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.120 min
Delta R.T.: 0.001 min
Response: 1313972
Conc: 57.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.117 min
Delta R.T.: 0.006 min
Response: 2452638
Conc: 52.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.530 min
Delta R.T.: 0.000 min
Response: 1080059
Conc: 51.97 ng/ml



#26 AR-1254-1


R.T.: 7.249 min
Delta R.T.: 0.002 min
Response: 1200825
Conc: 549.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.284 min
Delta R.T.: 0.000 min
Response: 740302
Conc: 545.92 ng/ml




#27 AR-1254-2

R.T.: 7.479 min
Delta R.T.: 0.003 min
Response: 1759210
Conc: 535.73 ng/ml



#27 AR-1254-2


R.T.: 6.445 min
Delta R.T.: 0.000 min
Response: 648643
Conc: 540.96 ng/ml



#28 AR-1254-3


R.T.: 7.862 min
Delta R.T.: 0.002 min
Response: 1754924
Conc: 509.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.872 min
Delta R.T.: 0.000 min
Response: 1019639
Conc: 544.95 ng/ml




#29 AR-1254-4

R.T.: 8.160 min
Delta R.T.: 0.002 min
Response: 1341743
Conc: 527.54 ng/ml



#29 AR-1254-4


R.T.: 7.114 min
Delta R.T.: 0.000 min
Response: 631727
Conc: 533.60 ng/ml



#30 AR-1254-5

R.T.: 8.590 min
Delta R.T.: 0.003 min
Response: 1335568
Conc: 502.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.549 min
Delta R.T.: 0.000 min
Response: 849759
Conc: 530.47 ng/ml