

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD 0\Data\P0073018\
 Data File : P0047396.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Jul 2018 23:43
 Operator : SM/SJ
 Sample : AR1268 MDL
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268 MDL

Manual Integrations
 APPROVED

Sohil
 7/31/2018 3:41:04 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 31 04:47:48 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 27 02:33:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.397	3.644	3882748	4114663	19.231	16.746
2) SA Decachlor...	10.087	8.636	3561556	3576977	21.837	22.755
Target Compounds						
41) L9 AR-1268-1	8.619	7.539	636988	793336	27.445	30.514
42) L9 AR-1268-2	8.712	7.603	568439	772324	26.530	31.001
43) L9 AR-1268-3	8.936	7.811	493297	634616	27.324	32.157
44) L9 AR-1268-4	9.354	8.094	222166	269405	27.862	32.116
45) L9 AR-1268-5	9.758	8.383	1533302	1692536	28.141m	31.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0073018\
Data File : P0047396.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Jul 2018 23:43
Operator : SM/SJ
Sample : AR1268 MDL
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1268 MDL


Manual Integrations
APPROVED

Sohil
7/31/2018 3:41:04 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 31 04:47:48 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0072718.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 27 02:33:29 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.397 min
Delta R.T.: 0.000 min
Response: 3882748
Conc: 19.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268 MDL


Manual Integrations
APPROVED

Sohil
7/31/2018 3:41:04 PM




#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: -0.002 min
Response: 4114663
Conc: 16.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.087 min
Delta R.T.: -0.003 min
Response: 3561556
Conc: 21.84 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.636 min
Delta R.T.: -0.008 min
Response: 3576977
Conc: 22.76 ng/ml




#41 AR-1268-1

R.T.: 8.619 min
Delta R.T.: -0.003 min
Response: 636988
Conc: 27.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268 MDL


Manual Integrations
APPROVED

Sohil
7/31/2018 3:41:04 PM




#41 AR-1268-1

R.T.: 7.539 min
Delta R.T.: -0.007 min
Response: 793336
Conc: 30.51 ng/ml




#42 AR-1268-2

R.T.: 8.712 min
Delta R.T.: -0.004 min
Response: 568439
Conc: 26.53 ng/ml



#42 AR-1268-2

R.T.: 7.603 min
Delta R.T.: -0.006 min
Response: 772324
Conc: 31.00 ng/ml




#43 AR-1268-3

R.T.: 8.936 min
Delta R.T.: -0.002 min
Response: 493297
Conc: 27.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268 MDL


Manual Integrations
APPROVED

Sohil
7/31/2018 3:41:04 PM




#43 AR-1268-3

R.T.: 7.811 min
Delta R.T.: -0.007 min
Response: 634616
Conc: 32.16 ng/ml




#44 AR-1268-4

R.T.: 9.354 min
Delta R.T.: -0.003 min
Response: 222166
Conc: 27.86 ng/ml



#44 AR-1268-4

R.T.: 8.094 min
Delta R.T.: -0.006 min
Response: 269405
Conc: 32.12 ng/ml




#45 AR-1268-5

R.T.: 9.758 min
Delta R.T.: -0.004 min
Response: 1533302
Conc: 28.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268 MDL

Manual Integrations
APPROVED

Sohil
7/31/2018 3:41:04 PM



#45 AR-1268-5

R.T.: 8.383 min
Delta R.T.: -0.007 min
Response: 1692536
Conc: 31.00 ng/ml