

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0073020\
 Data File : P0070198.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Jul 2020 5:31
 Operator : DD\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 06:31:57 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0073020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 30 06:08:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.376	3.559	2499465	2189113	50.000	50.000
2) SA Decachlor...	10.013	8.755	3535617	3268518	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.573	5.645	1289827	1626186	500.000	500.000
27) L6 AR-1254-2	6.801	5.806	2141963	1433306	500.000	500.000
28) L6 AR-1254-3	7.177	6.217	2239420	2299093	500.000	500.000
29) L6 AR-1254-4	7.472	6.457	2025127	1687926	500.000	500.000
30) L6 AR-1254-5	7.894	6.880	1936058	2240040	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO073020\
Data File : PO070198.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Jul 2020 5:31
Operator : DD\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 06:31:57 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO073020.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 30 06:08:47 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

