

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0073021\
 Data File : P0080011.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Jul 2021 14:52
 Operator : DD\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 16:26:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 28 18:45:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.000	4.118	3311057	1251102	54.948	54.631
2) SA Decachlor...	11.110	9.531	2476188	1082811	52.786	52.104
Target Compounds						
21) L5 AR-1248-1	6.331	5.396	667702	282247	526.710	511.133
22) L5 AR-1248-2	6.626	5.663	880729	406944	505.195	493.933
23) L5 AR-1248-3	6.845	5.709	1045569	422688	510.155	509.328
24) L5 AR-1248-4	7.276	5.897	1193989	474949	510.056	502.632
25) L5 AR-1248-5	7.315	6.326	1172835	487879	515.755	532.102



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO073021\
Data File : PO080011.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Jul 2021 14:52
Operator : DD\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 16:26:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072821.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 28 18:45:36 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.000 min
Delta R.T.: 0.000 min
Response: 3311057
Conc: 54.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.118 min
Delta R.T.: -0.001 min
Response: 1251102
Conc: 54.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.110 min
Delta R.T.: 0.000 min
Response: 2476188
Conc: 52.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.531 min
Delta R.T.: 0.002 min
Response: 1082811
Conc: 52.10 ng/ml



#21 AR-1248-1


R.T.: 6.331 min
Delta R.T.: -0.001 min
Response: 667702
Conc: 526.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.396 min
Delta R.T.: -0.002 min
Response: 282247
Conc: 511.13 ng/ml




#22 AR-1248-2

R.T.: 6.626 min
Delta R.T.: -0.002 min
Response: 880729
Conc: 505.19 ng/ml



#22 AR-1248-2


R.T.: 5.663 min
Delta R.T.: -0.002 min
Response: 406944
Conc: 493.93 ng/ml



#23 AR-1248-3


R.T.: 6.845 min
Delta R.T.: 0.000 min
Response: 1045569
Conc: 510.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.709 min
Delta R.T.: -0.002 min
Response: 422688
Conc: 509.33 ng/ml




#24 AR-1248-4

R.T.: 7.276 min
Delta R.T.: -0.001 min
Response: 1193989
Conc: 510.06 ng/ml



#24 AR-1248-4


R.T.: 5.897 min
Delta R.T.: -0.001 min
Response: 474949
Conc: 502.63 ng/ml



#25 AR-1248-5

R.T.: 7.315 min
Delta R.T.: -0.001 min
Response: 1172835
Conc: 515.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.326 min
Delta R.T.: 0.000 min
Response: 487879
Conc: 532.10 ng/ml