

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0073125\
 Data File : P0112632.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jul 2025 09:51
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 31 12:09:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072325.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 24 04:54:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.668	3.661	393.3E6	229.6E6	48.363	46.196
2) SA Decachlor...	8.692	8.637	378.3E6	86334952	51.740	48.857
Target Compounds						
26) L6 AR-1254-1	5.559	5.532	262.0E6	148.7E6	477.288	452.386
27) L6 AR-1254-2	5.708	5.680	237.3E6	132.2E6	488.305	456.624
28) L6 AR-1254-3	6.112	6.081	364.6E6	198.8E6	489.857	464.961
29) L6 AR-1254-4	6.341	6.309	269.4E6	127.4E6	490.051	466.734
30) L6 AR-1254-5	6.760	6.724	351.4E6	159.6E6	496.443	477.794



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO073125\
Data File : PO112632.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jul 2025 09:51
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 31 12:09:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072325.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 24 04:54:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.668 min
Delta R.T.: 0.000 min
Response: 393270482
Conc: 48.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.661 min
Delta R.T.: -0.002 min
Response: 229555714
Conc: 46.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.692 min
Delta R.T.: 0.000 min
Response: 378313650
Conc: 51.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.637 min
Delta R.T.: -0.003 min
Response: 86334952
Conc: 48.86 ng/ml



#26 AR-1254-1


R.T.: 5.559 min
Delta R.T.: -0.002 min
Response: 262011579
Conc: 477.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.532 min
Delta R.T.: -0.003 min
Response: 148663647
Conc: 452.39 ng/ml




#27 AR-1254-2

R.T.: 5.708 min
Delta R.T.: -0.002 min
Response: 237276696
Conc: 488.31 ng/ml



#27 AR-1254-2


R.T.: 5.680 min
Delta R.T.: -0.002 min
Response: 132188522
Conc: 456.62 ng/ml



#28 AR-1254-3


R.T.: 6.112 min
Delta R.T.: -0.001 min
Response: 364593717
Conc: 489.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.081 min
Delta R.T.: -0.003 min
Response: 198752885
Conc: 464.96 ng/ml




#29 AR-1254-4

R.T.: 6.341 min
Delta R.T.: -0.002 min
Response: 269383111
Conc: 490.05 ng/ml



#29 AR-1254-4


R.T.: 6.309 min
Delta R.T.: -0.003 min
Response: 127401163
Conc: 466.73 ng/ml



#30 AR-1254-5

R.T.: 6.760 min
Delta R.T.: -0.002 min
Response: 351394040
Conc: 496.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.724 min
Delta R.T.: -0.003 min
Response: 159592561
Conc: 477.79 ng/ml