

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080124\
 Data File : P0105118.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2024 22:11
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 02:25:17 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 30 07:11:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.429	3.540	402.0E6	147.2E6	58.400	57.775
2)	SA Decachlor...	10.221	8.432	260.8E6	80504534	50.792	50.679
Target Compounds							
21)	L5 AR-1248-1	5.600	4.595	82899534	32114500	534.737	524.038
22)	L5 AR-1248-2	5.875	4.826	116.2E6	42787537	526.639	514.759
23)	L5 AR-1248-3	6.082	4.866	128.0E6	45342326	521.521	520.297
24)	L5 AR-1248-4	6.488	5.035	145.2E6	53045753	480.095	514.010
25)	L5 AR-1248-5	6.527	5.419	140.5E6	52401786	494.331	509.668



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080124\
Data File : PO105118.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2024 22:11
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 02:25:17 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 30 07:11:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.429 min
Delta R.T.: 0.000 min
Response: 402047487
Conc: 58.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: 0.000 min
Response: 147214566
Conc: 57.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.221 min
Delta R.T.: 0.000 min
Response: 260762524
Conc: 50.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.432 min
Delta R.T.: -0.005 min
Response: 80504534
Conc: 50.68 ng/ml



#21 AR-1248-1


R.T.: 5.600 min
Delta R.T.: 0.001 min
Response: 82899534
Conc: 534.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.595 min
Delta R.T.: -0.001 min
Response: 32114500
Conc: 524.04 ng/ml




#22 AR-1248-2

R.T.: 5.875 min
Delta R.T.: 0.002 min
Response: 116171847
Conc: 526.64 ng/ml



#22 AR-1248-2


R.T.: 4.826 min
Delta R.T.: -0.001 min
Response: 42787537
Conc: 514.76 ng/ml



#23 AR-1248-3


R.T.: 6.082 min
Delta R.T.: 0.002 min
Response: 128044776
Conc: 521.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.866 min
Delta R.T.: -0.002 min
Response: 45342326
Conc: 520.30 ng/ml




#24 AR-1248-4

R.T.: 6.488 min
Delta R.T.: 0.002 min
Response: 145230592
Conc: 480.09 ng/ml



#24 AR-1248-4


R.T.: 5.035 min
Delta R.T.: 0.000 min
Response: 53045753
Conc: 514.01 ng/ml



#25 AR-1248-5

R.T.: 6.527 min
Delta R.T.: 0.002 min
Response: 140497155
Conc: 494.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.419 min
Delta R.T.: -0.002 min
Response: 52401786
Conc: 509.67 ng/ml