

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080222\
 Data File : P0088528.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2022 14:19
 Operator : YP/AJ
 Sample : N3966-07
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S3-03-S

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/03/2022
 Supervised By :Ankita Jodhani 08/03/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 00:01:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 26 11:20:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.443	3.603	54747016	19869704	21.459	21.060
2) SA Decachlor...	10.286	8.595	34884720	23186064	16.426	22.187 #
Target Compounds						
26) L6 AR-1254-1	6.478	5.478	43688495	25650758	460.785	444.901
27) L6 AR-1254-2	6.697	5.625	75987974	38325041	529.316	747.436 #
28) L6 AR-1254-3	7.067	6.028	150.3E6	85862660	1025.869	1058.078m
29) L6 AR-1254-4	7.354	6.256	94617562	50110149	874.245	1004.732
30) L6 AR-1254-5	7.782	6.674	209.3E6	571.0E6	1739.307m	7560.222 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080222\
Data File : PO088528.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2022 14:19
Operator : YP/AJ
Sample : N3966-07
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
S3-03-S


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/03/2022
Supervised By :Ankita Jodhani 08/03/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 03 00:01:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 26 11:20:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.443 min
Delta R.T.: -0.004 min
Response: 54747016
Conc: 21.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
S3-03-S


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/03/2022
Supervised By :Ankita Jodhani 08/03/2022




#1 Tetrachloro-m-xylene

R.T.: 3.603 min
Delta R.T.: -0.001 min
Response: 19869704
Conc: 21.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.286 min
Delta R.T.: -0.012 min
Response: 34884720
Conc: 16.43 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.595 min
Delta R.T.: -0.013 min
Response: 23186064
Conc: 22.19 ng/ml




#26 AR-1254-1

R.T.: 6.478 min
Delta R.T.: -0.007 min
Response: 43688495
Conc: 460.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
S3-03-S


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/03/2022
Supervised By :Ankita Jodhani 08/03/2022




#26 AR-1254-1

R.T.: 5.478 min
Delta R.T.: -0.008 min
Response: 25650758
Conc: 444.90 ng/ml




#27 AR-1254-2

R.T.: 6.697 min
Delta R.T.: -0.007 min
Response: 75987974
Conc: 529.32 ng/ml



#27 AR-1254-2

R.T.: 5.625 min
Delta R.T.: -0.009 min
Response: 38325041
Conc: 747.44 ng/ml




#28 AR-1254-3

R.T.: 7.067 min
Delta R.T.: -0.006 min
Response: 150265793
Conc: 1025.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
S3-03-S


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/03/2022
Supervised By :Ankita Jodhani 08/03/2022




#28 AR-1254-3

R.T.: 6.028 min
Delta R.T.: -0.009 min
Response: 85862660
Conc: 1058.08 ng/ml




#29 AR-1254-4

R.T.: 7.354 min
Delta R.T.: -0.005 min
Response: 94617562
Conc: 874.24 ng/ml



#29 AR-1254-4

R.T.: 6.256 min
Delta R.T.: -0.010 min
Response: 50110149
Conc: 1004.73 ng/ml




#30 AR-1254-5

R.T.: 7.782 min
Delta R.T.: 0.002 min
Response: 209331418
Conc: 1739.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
S3-03-S

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/03/2022
Supervised By :Ankita Jodhani 08/03/2022



#30 AR-1254-5

R.T.: 6.674 min
Delta R.T.: -0.009 min
Response: 570988994
Conc: 7560.22 ng/ml