

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080223\
 Data File : P0096587.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2023 14:43
 Operator : YP/AJ
 Sample : 03859-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 02:02:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 28 01:38:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.414	3.623	65316465	25894262	19.549	18.280
2) SA Decachlor...	10.241	8.646	46862285	23994896	21.180	20.398
Target Compounds						
3) L1 AR-1016-1	5.596	4.709	5656944	2852487	60.356	63.110
4) L1 AR-1016-2	5.619	4.728	7768847	3090772	58.229	49.204
5) L1 AR-1016-3	5.682	4.906	4008501	1759262	48.273	51.532
6) L1 AR-1016-4	5.779	4.947	3372681	773355	49.649	26.493 #
7) L1 AR-1016-5	6.079	5.160	778425	376874	10.872	9.914
8) L2 AR-1221-1	4.618	3.835	765269	346472	18.598	18.487
9) L2 AR-1221-2	4.715	3.923	2151480	1004389	71.574	70.918
10) L2 AR-1221-3	4.789	3.998	6617469	2766023	75.338	66.804
11) L3 AR-1232-1	4.789	3.998	6617469	2766023	92.006	81.414
12) L3 AR-1232-2	5.325	4.728	16002838	3090772	407.538	105.392 #
13) L3 AR-1232-3	5.619	4.906	7768847	1759262	127.473	112.871
14) L3 AR-1232-4	5.779	4.988	3372681	994156	111.141	65.427 #
15) L3 AR-1232-5	5.873	5.160	1893161	376874	68.055	22.625 #
16) L4 AR-1242-1	5.596	4.709	5656944	2852487	69.804	72.808
17) L4 AR-1242-2	5.619	4.728	7768847	3090772	67.900	57.485
18) L4 AR-1242-3	5.682	4.906	4008501	1759262	56.306	59.845
19) L4 AR-1242-4	5.779	4.988	3372681	994156	58.383	31.718 #
20) L4 AR-1242-5	6.529	5.511	40586	636391	0.732	17.640 #
21) L5 AR-1248-1	5.596	4.709	5656944	2852487	98.505	102.107
22) L5 AR-1248-2	5.873	4.947	1893161	773355	20.413	18.343
23) L5 AR-1248-3	6.079	4.988	778425	994156	7.900	22.626 #
24) L5 AR-1248-4	6.501	5.160	170828	376874	1.840	7.351 #
25) L5 AR-1248-5	6.529	5.592	40586	1001058	0.439	21.327 #
26) L6 AR-1254-1	6.461	5.511	98731	636391	0.934	8.660 #
27) L6 AR-1254-2	6.678	5.692	302331	800263	1.949	12.261 #
28) L6 AR-1254-3	7.047	6.063	1187787	517924	7.783	5.133 #
29) L6 AR-1254-4	7.338	6.309	81010	391188	0.829	6.558 #
30) L6 AR-1254-5	7.771	6.715	119962	251677	1.089	2.899 #
31) L7 AR-1260-1	7.218	6.237	419826	996118	3.227	13.669 #
32) L7 AR-1260-2	7.465	6.370	-59180	532877	N.D.	6.162 #
33) L7 AR-1260-3	7.755	0.000	1010023	0	6.606	N.D. #
34) L7 AR-1260-4	8.082	6.998	246788	299321	2.027	4.584 #
35) L7 AR-1260-5	8.390	7.229	158185	660736	0.719	4.663 #
36) L8 AR-1262-1	7.837	6.855f	594294	924960	4.025	23.544 #
37) L8 AR-1262-2	8.390	6.998	158185	299321	0.694	3.668 #
38) L8 AR-1262-3	8.727	7.595f	997722	482029	6.002	7.534 #
39) L8 AR-1262-4	8.806	7.595	474940	482029	3.673	4.255
40) L8 AR-1262-5	9.464	8.071f	144846	261918	1.698	5.033 #
41) L9 AR-1268-1	8.727	7.595f	997722	482029	3.280	2.583

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080223\
 Data File : P0096587.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2023 14:43
 Operator : YP/AJ
 Sample : 03859-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 02:02:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 28 01:38:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.806	7.595	474940	482029	1.774	2.865 #
43) L9	AR-1268-3	9.043	7.812	117394	250343	0.469	1.708 #
44) L9	AR-1268-4	9.464	8.071f	144846	261918	1.564	4.405 #
45) L9	AR-1268-5	9.902	8.393	77417	161472	0.097	0.374 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080223\
Data File : PO096587.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2023 14:43
Operator : YP/AJ
Sample : 03859-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 03 02:02:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072723.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 28 01:38:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.414 min
Delta R.T.: -0.004 min
Response: 65316465
Conc: 19.55 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.623 min
Delta R.T.: -0.004 min
Response: 25894262
Conc: 18.28 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.241 min
Delta R.T.: -0.011 min
Response: 46862285
Conc: 21.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.646 min
Delta R.T.: -0.009 min
Response: 23994896
Conc: 20.40 ng/ml



#3 AR-1016-1


R.T.: 5.596 min
Delta R.T.: -0.004 min
Response: 5656944
Conc: 60.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.709 min
Delta R.T.: -0.006 min
Response: 2852487
Conc: 63.11 ng/ml




#4 AR-1016-2

R.T.: 5.619 min
Delta R.T.: -0.004 min
Response: 7768847
Conc: 58.23 ng/ml



#4 AR-1016-2


R.T.: 4.728 min
Delta R.T.: -0.006 min
Response: 3090772
Conc: 49.20 ng/ml



#5 AR-1016-3


R.T.: 5.682 min
Delta R.T.: -0.003 min
Response: 4008501
Conc: 48.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.906 min
Delta R.T.: -0.004 min
Response: 1759262
Conc: 51.53 ng/ml




#6 AR-1016-4

R.T.: 5.779 min
Delta R.T.: -0.006 min
Response: 3372681
Conc: 49.65 ng/ml



#6 AR-1016-4


R.T.: 4.947 min
Delta R.T.: -0.006 min
Response: 773355
Conc: 26.49 ng/ml



#7 AR-1016-5


R.T.: 6.079 min
Delta R.T.: -0.004 min
Response: 778425
Conc: 10.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.160 min
Delta R.T.: -0.006 min
Response: 376874
Conc: 9.91 ng/ml




#8 AR-1221-1

R.T.: 4.618 min
Delta R.T.: -0.011 min
Response: 765269
Conc: 18.60 ng/ml



#8 AR-1221-1


R.T.: 3.835 min
Delta R.T.: -0.007 min
Response: 346472
Conc: 18.49 ng/ml



#9 AR-1221-2


R.T.: 4.715 min
Delta R.T.: -0.002 min
Response: 2151480
Conc: 71.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.923 min
Delta R.T.: -0.004 min
Response: 1004389
Conc: 70.92 ng/ml




#10 AR-1221-3

R.T.: 4.789 min
Delta R.T.: -0.005 min
Response: 6617469
Conc: 75.34 ng/ml



#10 AR-1221-3


R.T.: 3.998 min
Delta R.T.: -0.005 min
Response: 2766023
Conc: 66.80 ng/ml



#11 AR-1232-1


R.T.: 4.789 min
Delta R.T.: -0.004 min
Response: 6617469
Conc: 92.01 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.998 min
Delta R.T.: -0.005 min
Response: 2766023
Conc: 81.41 ng/ml




#12 AR-1232-2

R.T.: 5.325 min
Delta R.T.: -0.005 min
Response: 16002838
Conc: 407.54 ng/ml



#12 AR-1232-2


R.T.: 4.728 min
Delta R.T.: -0.006 min
Response: 3090772
Conc: 105.39 ng/ml



#13 AR-1232-3


R.T.: 5.619 min
Delta R.T.: -0.004 min
Response: 7768847
Conc: 127.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.906 min
Delta R.T.: -0.004 min
Response: 1759262
Conc: 112.87 ng/ml




#14 AR-1232-4

R.T.: 5.779 min
Delta R.T.: -0.007 min
Response: 3372681
Conc: 111.14 ng/ml



#14 AR-1232-4


R.T.: 4.988 min
Delta R.T.: -0.007 min
Response: 994156
Conc: 65.43 ng/ml



#15 AR-1232-5


R.T.: 5.873 min
Delta R.T.: -0.005 min
Response: 1893161
Conc: 68.05 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.160 min
Delta R.T.: -0.006 min
Response: 376874
Conc: 22.63 ng/ml




#16 AR-1242-1

R.T.: 5.596 min
Delta R.T.: -0.005 min
Response: 5656944
Conc: 69.80 ng/ml



#16 AR-1242-1


R.T.: 4.709 min
Delta R.T.: -0.006 min
Response: 2852487
Conc: 72.81 ng/ml



#17 AR-1242-2


R.T.: 5.619 min
Delta R.T.: -0.005 min
Response: 7768847
Conc: 67.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.728 min
Delta R.T.: -0.006 min
Response: 3090772
Conc: 57.48 ng/ml




#18 AR-1242-3

R.T.: 5.682 min
Delta R.T.: -0.004 min
Response: 4008501
Conc: 56.31 ng/ml



#18 AR-1242-3


R.T.: 4.906 min
Delta R.T.: -0.004 min
Response: 1759262
Conc: 59.84 ng/ml



#19 AR-1242-4


R.T.: 5.779 min
Delta R.T.: -0.007 min
Response: 3372681
Conc: 58.38 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.988 min
Delta R.T.: -0.007 min
Response: 994156
Conc: 31.72 ng/ml




#20 AR-1242-5

R.T.: 6.529 min
Delta R.T.: 0.000 min
Response: 40586
Conc: 0.73 ng/ml



#20 AR-1242-5


R.T.: 5.511 min
Delta R.T.: -0.009 min
Response: 636391
Conc: 17.64 ng/ml



#21 AR-1248-1


R.T.: 5.596 min
Delta R.T.: -0.005 min
Response: 5656944
Conc: 98.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.709 min
Delta R.T.: -0.006 min
Response: 2852487
Conc: 102.11 ng/ml




#22 AR-1248-2

R.T.: 5.873 min
Delta R.T.: -0.004 min
Response: 1893161
Conc: 20.41 ng/ml



#22 AR-1248-2


R.T.: 4.947 min
Delta R.T.: -0.006 min
Response: 773355
Conc: 18.34 ng/ml



#23 AR-1248-3


R.T.: 6.079 min
Delta R.T.: -0.006 min
Response: 778425
Conc: 7.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.988 min
Delta R.T.: -0.007 min
Response: 994156
Conc: 22.63 ng/ml




#24 AR-1248-4

R.T.: 6.501 min
Delta R.T.: 0.010 min
Response: 170828
Conc: 1.84 ng/ml



#24 AR-1248-4


R.T.: 5.160 min
Delta R.T.: -0.007 min
Response: 376874
Conc: 7.35 ng/ml



#25 AR-1248-5


R.T.: 6.529 min
Delta R.T.: -0.001 min
Response: 40586
Conc: 0.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.592 min
Delta R.T.: 0.030 min
Response: 1001058
Conc: 21.33 ng/ml




#26 AR-1254-1

R.T.: 6.461 min
Delta R.T.: -0.005 min
Response: 98731
Conc: 0.93 ng/ml



#26 AR-1254-1


R.T.: 5.511 min
Delta R.T.: -0.010 min
Response: 636391
Conc: 8.66 ng/ml



#27 AR-1254-2


R.T.: 6.678 min
Delta R.T.: -0.008 min
Response: 302331
Conc: 1.95 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.692 min
Delta R.T.: 0.023 min
Response: 800263
Conc: 12.26 ng/ml




#28 AR-1254-3

R.T.: 7.047 min
Delta R.T.: -0.007 min
Response: 1187787
Conc: 7.78 ng/ml



#28 AR-1254-3


R.T.: 6.063 min
Delta R.T.: -0.012 min
Response: 517924
Conc: 5.13 ng/ml



#29 AR-1254-4


R.T.: 7.338 min
Delta R.T.: -0.004 min
Response: 81010
Conc: 0.83 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.309 min
Delta R.T.: 0.005 min
Response: 391188
Conc: 6.56 ng/ml




#30 AR-1254-5

R.T.: 7.771 min
Delta R.T.: 0.006 min
Response: 119962
Conc: 1.09 ng/ml



#30 AR-1254-5


R.T.: 6.715 min
Delta R.T.: -0.008 min
Response: 251677
Conc: 2.90 ng/ml



#31 AR-1260-1


R.T.: 7.218 min
Delta R.T.: -0.002 min
Response: 419826
Conc: 3.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.237 min
Delta R.T.: 0.033 min
Response: 996118
Conc: 13.67 ng/ml




#32 AR-1260-2

R.T.: 7.465 min
Delta R.T.: -0.012 min
Response: -59180
Conc: N.D.



#32 AR-1260-2


R.T.: 6.370 min
Delta R.T.: -0.023 min
Response: 532877
Conc: 6.16 ng/ml



#33 AR-1260-3


R.T.: 7.755 min
Delta R.T.: -0.008 min
Response: 1010023
Conc: 6.61 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T.: 6.547 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 8.082 min
Delta R.T.: 0.016 min
Response: 246788
Conc: 2.03 ng/ml



#34 AR-1260-4


R.T.: 6.998 min
Delta R.T.: -0.022 min
Response: 299321
Conc: 4.58 ng/ml



#35 AR-1260-5


R.T.: 8.390 min
Delta R.T.: -0.002 min
Response: 158185
Conc: 0.72 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.229 min
Delta R.T.: -0.034 min
Response: 660736
Conc: 4.66 ng/ml




#36 AR-1262-1

R.T.: 7.837 min
Delta R.T.: -0.003 min
Response: 594294
Conc: 4.03 ng/ml



#36 AR-1262-1


R.T.: 6.855 min
Delta R.T.: 0.041 min
Response: 924960
Conc: 23.54 ng/ml



#37 AR-1262-2


R.T.: 8.390 min
Delta R.T.: -0.002 min
Response: 158185
Conc: 0.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 6.998 min
Delta R.T.: -0.023 min
Response: 299321
Conc: 3.67 ng/ml




#38 AR-1262-3

R.T.: 8.727 min
Delta R.T.: 0.016 min
Response: 997722
Conc: 6.00 ng/ml



#38 AR-1262-3


R.T.: 7.595 min
Delta R.T.: 0.046 min
Response: 482029
Conc: 7.53 ng/ml



#39 AR-1262-4


R.T.: 8.806 min
Delta R.T.: 0.005 min
Response: 474940
Conc: 3.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.595 min
Delta R.T.: -0.018 min
Response: 482029
Conc: 4.26 ng/ml




#40 AR-1262-5

R.T.: 9.464 min
Delta R.T.: -0.010 min
Response: 144846
Conc: 1.70 ng/ml



#40 AR-1262-5


R.T.: 8.071 min
Delta R.T.: -0.039 min
Response: 261918
Conc: 5.03 ng/ml



#41 AR-1268-1


R.T.: 8.727 min
Delta R.T.: 0.018 min
Response: 997722
Conc: 3.28 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.595 min
Delta R.T.: 0.046 min
Response: 482029
Conc: 2.58 ng/ml




#42 AR-1268-2

R.T.: 8.806 min
Delta R.T.: 0.001 min
Response: 474940
Conc: 1.77 ng/ml



#42 AR-1268-2


R.T.: 7.595 min
Delta R.T.: -0.019 min
Response: 482029
Conc: 2.86 ng/ml



#43 AR-1268-3


R.T.: 9.043 min
Delta R.T.: 0.004 min
Response: 117394
Conc: 0.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.812 min
Delta R.T.: -0.009 min
Response: 250343
Conc: 1.71 ng/ml




#44 AR-1268-4

R.T.: 9.464 min
Delta R.T.: -0.011 min
Response: 144846
Conc: 1.56 ng/ml



#44 AR-1268-4


R.T.: 8.071 min
Delta R.T.: -0.039 min
Response: 261918
Conc: 4.40 ng/ml



#45 AR-1268-5

R.T.: 9.902 min
Delta R.T.: 0.000 min
Response: 77417
Conc: 0.10 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.393 min
Delta R.T.: -0.008 min
Response: 161472
Conc: 0.37 ng/ml