

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080223\
 Data File : P0096613.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2023 22:53
 Operator : YP/AJ
 Sample : 03862-08
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BUILDING-B-1-(0-5)

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/14/2023
 Supervised By :Ankita Jodhani 08/14/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 02:10:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 28 01:38:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.408	3.622	67148889	24863126	20.097	17.552
2) SA Decachlor...	10.229	8.643	46016112	22793768	20.798m	19.377m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080223\
Data File : PO096613.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2023 22:53
Operator : YP/AJ
Sample : 03862-08
Misc :
ALS Vial : 28 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
BUILDING-B-1-(0-5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2023
Supervised By :Ankita Jodhani 08/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 03 02:10:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072723.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 28 01:38:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.408 min
Delta R.T.: -0.010 min
Response: 67148889
Conc: 20.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
BUILDING-B-1-(0-5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2023
Supervised By :Ankita Jodhani 08/14/2023




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: -0.005 min
Response: 24863126
Conc: 17.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.229 min
Delta R.T.: -0.022 min
Response: 46016112
Conc: 20.80 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.643 min
Delta R.T.: -0.012 min
Response: 22793768
Conc: 19.38 ng/ml m