

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080319\
 Data File : P0058843.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Aug 2019 1:23
 Operator : SM/AJ
 Sample : AR1660CCC500

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Misc :
 ALS Vial : 3 Sample Multiplier: 1

Manual Integrations
 APPROVED

Ankita
 8/5/2019 11:14:49 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 01:38:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072819.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 27 06:49:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.161	3.516	2363716	1996013	56.009	49.963
2)	SA Decachlor...	9.679	8.398	2371512	1831730	43.791	46.405
Target Compounds							
3)	L1 AR-1016-1	5.318	4.572	974808	708286	520.189m	485.799m
4)	L1 AR-1016-2	5.339	4.590	1375145	991553	499.773m	461.918
5)	L1 AR-1016-3	5.401	4.763	836052	521440	492.394	459.123
6)	L1 AR-1016-4	5.499	4.804	716880	448686	512.349	475.654
7)	L1 AR-1016-5	5.788	5.013	744526	572239	519.763	464.692
31)	L7 AR-1260-1	6.901	6.025	1425627	1039215	495.156	436.636
32)	L7 AR-1260-2	7.157	6.213	1886680	1310873	453.523	432.731
33)	L7 AR-1260-3	7.512	6.362	1631862	1142897	438.205	423.320
34)	L7 AR-1260-4	7.739	6.826	1496018	995168	439.001	440.432
35)	L7 AR-1260-5	8.047	7.069	3200695	2365023	435.187	439.159

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080319\
Data File : PO058843.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Aug 2019 1:23
Operator : SM/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
8/5/2019 11:14:49 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 03 01:38:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072819.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 27 06:49:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.161 min
Delta R.T.: -0.018 min
Response: 2363716
Conc: 56.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
8/5/2019 11:14:49 AM




#1 Tetrachloro-m-xylene

R.T.: 3.516 min
Delta R.T.: -0.008 min
Response: 1996013
Conc: 49.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.679 min
Delta R.T.: 0.006 min
Response: 2371512
Conc: 43.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.398 min
Delta R.T.: -0.016 min
Response: 1831730
Conc: 46.40 ng/ml




#3 AR-1016-1

R.T.: 5.318 min
Delta R.T.: -0.017 min
Response: 974808
Conc: 520.19 ng/ml m

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
8/5/2019 11:14:49 AM




#3 AR-1016-1

R.T.: 4.572 min
Delta R.T.: -0.011 min
Response: 708286
Conc: 485.80 ng/ml m




#4 AR-1016-2

R.T.: 5.339 min
Delta R.T.: -0.018 min
Response: 1375145
Conc: 499.77 ng/ml m



#4 AR-1016-2

R.T.: 4.590 min
Delta R.T.: -0.011 min
Response: 991553
Conc: 461.92 ng/ml




#5 AR-1016-3

R.T.: 5.401 min
Delta R.T.: -0.016 min
Response: 836052
Conc: 492.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
8/5/2019 11:14:49 AM




#5 AR-1016-3

R.T.: 4.763 min
Delta R.T.: -0.011 min
Response: 521440
Conc: 459.12 ng/ml




#6 AR-1016-4

R.T.: 5.499 min
Delta R.T.: -0.016 min
Response: 716880
Conc: 512.35 ng/ml



#6 AR-1016-4

R.T.: 4.804 min
Delta R.T.: -0.011 min
Response: 448686
Conc: 475.65 ng/ml




#7 AR-1016-5

R.T.: 5.788 min
Delta R.T.: -0.015 min
Response: 744526
Conc: 519.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
8/5/2019 11:14:49 AM




#7 AR-1016-5

R.T.: 5.013 min
Delta R.T.: -0.011 min
Response: 572239
Conc: 464.69 ng/ml




#31 AR-1260-1

R.T.: 6.901 min
Delta R.T.: -0.011 min
Response: 1425627
Conc: 495.16 ng/ml



#31 AR-1260-1

R.T.: 6.025 min
Delta R.T.: -0.012 min
Response: 1039215
Conc: 436.64 ng/ml




#32 AR-1260-2

R.T.: 7.157 min
Delta R.T.: -0.010 min
Response: 1886680
Conc: 453.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
8/5/2019 11:14:49 AM




#32 AR-1260-2

R.T.: 6.213 min
Delta R.T.: -0.011 min
Response: 1310873
Conc: 432.73 ng/ml




#33 AR-1260-3

R.T.: 7.512 min
Delta R.T.: -0.008 min
Response: 1631862
Conc: 438.21 ng/ml



#33 AR-1260-3

R.T.: 6.362 min
Delta R.T.: -0.012 min
Response: 1142897
Conc: 423.32 ng/ml




#34 AR-1260-4

R.T.: 7.739 min
Delta R.T.: -0.007 min
Response: 1496018
Conc: 439.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
8/5/2019 11:14:49 AM




#34 AR-1260-4

R.T.: 6.826 min
Delta R.T.: -0.014 min
Response: 995168
Conc: 440.43 ng/ml



#35 AR-1260-5

R.T.: 8.047 min
Delta R.T.: -0.004 min
Response: 3200695
Conc: 435.19 ng/ml



#35 AR-1260-5

R.T.: 7.069 min
Delta R.T.: -0.013 min
Response: 2365023
Conc: 439.16 ng/ml