

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0080519\
 Data File : P0058877.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Aug 2019 19:10
 Operator : SM/AJ
 Sample : K4153-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 RT-4221

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 01:11:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072819.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 27 06:49:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.158	3.515	1081915	884972	25.636	22.152
2) SA Decachlor...	9.666	8.399	774136	601052	14.295	15.227
Target Compounds						
10) L2 AR-1221-3	0.000	3.925	0	23177	N.D.	24.228 #
11) L3 AR-1232-1	0.000	3.925	0	23177	N.D.	29.504 #
30) L6 AR-1254-5	0.000	6.534	0	20513	N.D.	5.947 #
32) L7 AR-1260-2	0.000	6.286	0	38587	N.D.	12.738 #
36) L8 AR-1262-1	0.000	6.534	0	20513	N.D.	5.743 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO080519\
Data File : PO058877.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Aug 2019 19:10
Operator : SM/AJ
Sample : K4153-01
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
RT-4221

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 01:11:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072819.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 27 06:49:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.158 min
Delta R.T.: -0.021 min
Response: 1081915
Conc: 25.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
RT-4221




#1 Tetrachloro-m-xylene

R.T.: 3.515 min
Delta R.T.: -0.010 min
Response: 884972
Conc: 22.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.666 min
Delta R.T.: -0.007 min
Response: 774136
Conc: 14.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.399 min
Delta R.T.: -0.015 min
Response: 601052
Conc: 15.23 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 4.553 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
RT-4221




#10 AR-1221-3

R.T.: 3.925 min
Delta R.T.: 0.035 min
Response: 23177
Conc: 24.23 ng/ml




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.554 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 3.925 min
Delta R.T.: 0.035 min
Response: 23177
Conc: 29.50 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 7.450 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
RT-4221




#30 AR-1254-5

R.T.: 6.534 min
Delta R.T.: -0.014 min
Response: 20513
Conc: 5.95 ng/ml




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.167 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 6.286 min
Delta R.T.: 0.061 min
Response: 38587
Conc: 12.74 ng/ml



#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 7.520 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
RT-4221



#36 AR-1262-1

R.T.: 6.534 min
Delta R.T.: -0.049 min
Response: 20513
Conc: 5.74 ng/ml